



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

**DATAMASTER MAINTENANCE REPORT**

RECEIVED 7/10/14 CD

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

**REVIEWED**  
By Carol Day at 3:37 pm, Aug 12, 2014

|   |   |                                  |
|---|---|----------------------------------|
| DATAMASTER SN<br>204106   | NAME OF AGENCY<br>Missouri State Highway Patrol | DATE OF INSPECTION<br>06/28/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>Monroe City Police Department, 300 North Main Street, Monroe City, MO |   | TIME OF INSPECTION<br>8:47 pm    |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

|  |   |
|--|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) | DATE AND TIME (from printout) 06/28/2014 20:47      |
| <input checked="" type="checkbox"/> COMPUTER                             | <input checked="" type="checkbox"/> DETECTOR        |
| <input checked="" type="checkbox"/> PROGRAM                              | <input checked="" type="checkbox"/> FILTERS         |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C         | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR                        | <input checked="" type="checkbox"/> CALIBRATION     |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED                      | <input checked="" type="checkbox"/> PRINTER         |

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 13290 EXP. DATE 10/29/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN G6817 EXP. DATE 05/02/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

|              |              |              |
|--------------|--------------|--------------|
| TEST 1  .097 | TEST 2  .097 | TEST 3  .098 |
|--------------|--------------|--------------|

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

|          |   |         |   |           |   |           |   |           |   |          |   |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|----------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | OVER .19 | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|----------|---|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

|  |   |
|--|---|
| <b>INSPECTING OFFICER</b>                                  |   |
| SIGNATURE<br>  | PRINT FULL NAME<br>James Johnson, Trooper |
| TYPE II PERMIT NUMBER/EXPIRATION DATE<br>240173 04/22/2016 | TELEPHONE NUMBER<br>(660) 385-2132        |

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-584-6470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**JAMES P JOHNSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2014

NUMBER 240173

EXPIRES 4/22/2016

580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator JOHNSON, JAMES  
Permit No 240173  
Date Issued 4/22/2014 Date Expires 4/22/2016

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204106  
06/28/14

TESTING OFFICER:  
JOHNSON, JAMES P  
OFFICER I.D.#: 487  
PERMIT NUMBER: 240173  
EXPIRATION DATE: 04/22/16  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

|                   |          |       |
|-------------------|----------|-------|
| BLANK TEST        | .000     | 20:50 |
| INTERNAL STANDARD | VERIFIED | 20:50 |
| EXTERNAL STANDARD | .097     | 20:50 |
| BLANK TEST        | .000     | 20:51 |
| EXTERNAL STANDARD | .097     | 20:51 |
| BLANK TEST        | .000     | 20:52 |
| EXTERNAL STANDARD | .099     | 20:52 |
| BLANK TEST        | .000     | 20:53 |

N = 3  
STDEV = .1  
AVG = .0973

Operator Signature

*[Signature]* #487

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204106  
06/28/14

ARRIVAL TIME: 20:00  
SUBJECT NAME:  
RFI/TEST  
DOB: (1/11/11) SEX: M  
STATE/D.L.#: NA/NA  
ARRESTING OFFICER:  
NA  
OFFICER I.D.#: NA  
TESTING OFFICER:  
JOHNSON, JAMES P  
OFFICER I.D.#: 487  
PERMIT NUMBER: 240173  
EXPIRATION DATE: 04/22/16  
MISCELLANEOUS DATA:  
RADIO FREQUENCY INTERFERENCE TEST

--- BREATH ANALYSIS ---

|                    |          |       |
|--------------------|----------|-------|
| BLANK TEST         | .000     | 20:56 |
| INTERNAL STANDARD  | VERIFIED | 20:57 |
| RADIO INTERFERENCE |          |       |

Operator Signature

*[Signature]* #487

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATA MASTER SERIAL NUMBER 264106  
06/28/14  
20:47

--- DIAGNOSTIC CHECK ---

|                       |      |
|-----------------------|------|
| COMPUTER:             | OKAY |
| PROGRAM (04-07-2009): | OKAY |
| HEATERS               |      |
| SAMPLE CHAMBER:       | 56c  |
| FLOW DETECTOR:        | OKAY |
| PUMP                  |      |
| HIGH SPEED:           | OKAY |
| DETECTOR:             | OKAY |
| FILTERS:              | OKAY |
| QUARTZ STANDARD:      | OKAY |
| CALIBRATION:          | OKAY |

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~

Operator Signature

*[Handwritten Signature]* #487