



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

received 3/9/14-cd  
 REPORT #6

**REVIEWED**  
 By Carol Day at 1:45 pm, Mar 19, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

|   |   |                                  |
|---|---|----------------------------------|
| DATAMASTER SN<br>204113   | NAME OF AGENCY<br>Missouri State Highway Patrol | DATE OF INSPECTION<br>02/28/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>Carter County Sheriff's office, Van Buren, MO 63965 |   | TIME OF INSPECTION<br>8:34 am    |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

|  |   |
|--|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) | DATE AND TIME (from printout) 02/28/2014 0834       |
| <input checked="" type="checkbox"/> COMPUTER                             | <input checked="" type="checkbox"/> DETECTOR        |
| <input checked="" type="checkbox"/> PROGRAM                              | <input checked="" type="checkbox"/> FILTERS         |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49°C    | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR                        | <input checked="" type="checkbox"/> CALIBRATION     |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED                      | <input checked="" type="checkbox"/> PRINTER         |

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN G11070 EXP. DATE 12/31/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

|               |               |               |
|---------------|---------------|---------------|
| TEST 1 ← .104 | TEST 2 ← .105 | TEST 3 ← .105 |
|---------------|---------------|---------------|

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |   |         |   |           |   |           |   |           |   |          |   |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|----------|---|
| REFUSALS | 0 | (0-.04) | 1 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | OVER .19 | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|----------|---|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

|  |   |
|--|---|
| <b>INSPECTING OFFICER</b>                                  |   |
| SIGNATURE<br>Sgt C.N Ponder 600                            | PRINT FULL NAME<br>Sgt. C. N. Ponder #600 |
| TYPE II PERMIT NUMBER/EXPIRATION DATE<br>220304 09/25/2014 | TELEPHONE NUMBER<br>(417) 469-3121        |

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



CRAIG N PONDER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/25/2012

Number 220304

Expires 09/25/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204113  
02/28/14  
08:34

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 49c  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~

Operator Signature

*SGT C N Pouch 600*

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204113  
02/28/14

TESTING OFFICER:  
PONDER/C/N  
OFFICER I.D.: 600  
PERMIT NUMBER: 220304  
EXPIRATION DATE: 09/25/14  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

|                   |          |       |
|-------------------|----------|-------|
| BLANK TEST        | .000     | 08:36 |
| INTERNAL STANDARD | VERIFIED | 08:36 |
| EXTERNAL STANDARD | .104     | 08:36 |
| BLANK TEST        | .000     | 08:37 |
| EXTERNAL STANDARD | .105     | 08:37 |
| BLANK TEST        | .000     | 08:38 |
| EXTERNAL STANDARD | .105     | 08:38 |
| BLANK TEST        | .000     | 08:39 |

N = 3  
SIM. = .1  
AVG. = .1046

Operator Signature

*SGT C N Pouch 600*

Face This Side Down - This Edge In First

**BAC DataMaster**  
**Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204113  
02/28/14

ARREST TIME: 08:00  
SUBJECT NAME:  
TEST  
DOB: 01/01/50      SEX: M  
STATE/D.L.: MO/12345  
ARRESTING OFFICER:  
PONDER/C/N  
OFFICER I.D.: 600  
TESTING OFFICER:  
SAME  
OFFICER I.D.:  
PERMIT NUMBER: 220304  
EXPIRATION DATE: 09/25/14  
MISCELLANEOUS DATA:  
RFI TEST

--- BREATH ANALYSIS ---

|                    |          |       |
|--------------------|----------|-------|
| BLANK TEST         | .000     | 08:42 |
| INTERNAL STANDARD  | VERIFIED | 08:42 |
| RADIO INTERFERENCE |          |       |

Operator Signature

SGT C N Ponder 600