



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #
 By Carol Day at 8:51 am, Feb 11, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204114	NAME OF AGENCY MISSOURI STATE HIGHWAY PATROL - TROOP G	DATE OF INSPECTION 01/30/2014
LOCATION OF INSTRUMENT (STREET AND CITY) OZARK COUNTY SHERIFF'S OFFICE, HC1 BOX 8-1, GAINESVILLE, MO		TIME OF INSPECTION 0913

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01/30/2014</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER REPCO MARKETING, INC LOT # 12002 EXP. DATE 08/29/2014

SIMULATOR TEMP (34°C ± 0.2°C) 34.1°C °C SIMULATOR SN G11072 EXP. DATE 11-19-2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 = .103% TEST 2 = .104% TEST 3 = .103%

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>0</u>	OVER .19 <u>0</u>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within the guidelines set by the Missouri Department of Health. RepCo Marketing Inc., .100% solution, Lot #12002, bottle #0739, expiration date 08-29-2014.

INSPECTING OFFICER

SIGNATURE <u>Sgt. Charley A. Hogue</u>	PRINT FULL NAME Sgt. Charley A. Hogue
TYPE II PERMIT NUMBER/EXPIRATION DATE 220313 09/28/2014	TELEPHONE NUMBER (417) 469-3121

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 12002
EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
T Y P E I I



CHARLEY A HOGUE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220313

Expires 09/28/2014

MO 560-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (F7-88)

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204114
01/30/14

ARREST TIME: 09:00
SUBJECT NAME:
SUBJECT/A/
DOB: 11/15/68 SEX: F
STATE/D.L.: AR/9876543210
ARRESTING OFFICER:
HOGUE/CHARLEY/A
OFFICER I.D.: 189
TESTING OFFICER:
HOGUE/CHARLEY/A
OFFICER I.D.: 189
PERMIT NUMBER: 220313
EXPIRATION DATE: 09/28/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 09:26
INTERNAL STANDARD VERIFIED 09:26
RADIO INTERFERENCE

Operator Signature

SGT. C.A. Hogue

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204114
01/30/14

ARREST TIME: 09:00
SUBJECT NAME:
SUBJECT/A/
DOB: 11/15/68 SEX: F
STATE/D.L.: AR/9876543210
ARRESTING OFFICER:
HOGUE/CHARLEY/A
OFFICER I.D.: 189
TESTING OFFICER:
HOGUE/CHARLEY/A
OFFICER I.D.: 189
PERMIT NUMBER: 220313
EXPIRATION DATE: 09/28/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 09:29
INTERNAL STANDARD VERIFIED 09:29
SUBJECT SAMPLE .000 09:31
BLANK TEST .000 09:31

Operator Signature

SGT. C.A. Hogue

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204114
01/30/14
09:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 50c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopq
rstuvwxyz{|}~+

OPERATOR SIGNATURE

SGT. C.A. Hogue

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204114
01/30/14

TESTING OFFICER:
HOGUE/CHARLEY/A
OFFICER I.D.: 189
PERMIT NUMBER: 220313
EXPIRATION DATE: 09/25/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	09:18
INTERNAL STANDARD	VERIFIED	09:18
EXTERNAL STANDARD	.103	09:18
BLANK TEST	.000	09:19
EXTERNAL STANDARD	.104	09:19
BLANK TEST	.000	09:20
EXTERNAL STANDARD	.103	09:20
BLANK TEST	.000	09:21

H = 3
SIN. = .1
AVG. = .1033

Operator Signature

SGT. C.A. Hogue