



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 12:48 pm, Jul 03, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204127 Inv#127256	NAME OF AGENCY MSHP	DATE OF INSPECTION 07/02/2014
LOCATION OF INSTRUMENT (STREET AND CITY) #5 Basler Drive, Sainte Genevieve, MO 63670		TIME OF INSPECTION 9:51 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07-02-2014 2151 hours
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER REPCO MARKETING Inc. LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.1 °C SIMULATOR SN G6828 EXP. DATE 07/09/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .103	TEST 2 .103	TEST 3 .103
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	11	(.05-.09)	1	(.10-.14)	3	(.15-.19)	3	OVER .19	0
----------	---	---------	----	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
 (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE Cpl. B. J. Fowler	PRINT FULL NAME Cpl. B. J. Fowler
--------------------------------	--------------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 240050 03/07/2016	TELEPHONE NUMBER (636) 300-2800
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

1. DATE OF COLLECTION: 01/15/2015
2. TIME OF COLLECTION: 10:00 AM
3. LOCATION OF COLLECTION: 12345 Main St, Anytown, CA 90210
4. NAME OF OPERATOR: [Signature]
5. NAME OF EVIDENCE: BAC DataMaster
6. QUANTITY OF EVIDENCE: 1
7. WEIGHT OF EVIDENCE: 1.2 g
8. VOLUME OF EVIDENCE: 0.5 mL
9. TEMPERATURE OF EVIDENCE: 20°C
10. OTHER INFORMATION: None

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

1. DATE OF COLLECTION: 01/15/2015
2. TIME OF COLLECTION: 10:00 AM
3. LOCATION OF COLLECTION: 12345 Main St, Anytown, CA 90210
4. NAME OF OPERATOR: [Signature]
5. NAME OF EVIDENCE: BAC DataMaster
6. QUANTITY OF EVIDENCE: 1
7. WEIGHT OF EVIDENCE: 1.2 g
8. VOLUME OF EVIDENCE: 0.5 mL
9. TEMPERATURE OF EVIDENCE: 20°C
10. OTHER INFORMATION: None

Operator Signature




Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

1. DATE OF COLLECTION: 01/15/2015
2. TIME OF COLLECTION: 10:00 AM
3. LOCATION OF COLLECTION: 12345 Main St, Anytown, CA 90210
4. NAME OF OPERATOR: [Signature]
5. NAME OF EVIDENCE: BAC DataMaster
6. QUANTITY OF EVIDENCE: 1
7. WEIGHT OF EVIDENCE: 1.2 g
8. VOLUME OF EVIDENCE: 0.5 mL
9. TEMPERATURE OF EVIDENCE: 20°C
10. OTHER INFORMATION: None

Operator Signature





STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRENT J FOWLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.

DATE 3/7/2014

NUMBER 240050

EXPIRES 3/7/2016

MO 680-0771 (0-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-1 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **FOWLER, BRENT**
Permit No **240050**
Date Issued **3/7/2014** Date Expires **3/7/2016**