



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 1/28/14-CD REPORT #

REVIEWED (days)
 By Carol Day at 12:24 pm, Feb 04, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204131	NAME OF AGENCY MSHP H-10	DATE OF INSPECTION 01-14-2014
--------------------------------	------------------------------------	---

LOCATION OF INSTRUMENT (STREET AND CITY) DEKALB CO. SHERIFF'S OPT MAYSVILLE, MO.	TIME OF INSPECTION 2212
--	-----------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed value where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01-14-14 / 2212
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER REPCO MARKETING INC LOT # 12002 EXP. DATE 08-29-14
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 33² °C SIMULATOR SN G6786 EXP. DATE 07-09-14

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.097	TEST 2 0.099	TEST 3 0.099
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 1	(.05-.09) 0	(.10-.14) 1	(.15-.19) 1	OVER .19 0
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
N/A

INSPECTING OFFICER	
SIGNATURE <i>[Signature]</i>	PRINT FULL NAME P.H. KIMBALL
TYPE II PERMIT NUMBER/EXPIRATION DATE 220235 09-07-14	TELEPHONE NUMBER 616-387-2345

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd.**

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



PAUL M KIMBALL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/07/2012

Number 220235

Expires 09/07/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204131
01/14/14
22:12

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 50c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJK
LMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstu
vwxyz{|}~

Operator Signature

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204131
01/14/14

TESTING OFFICER:
KIMBALL
OFFICER I.D.: 1149
PERMIT NUMBER: 220235
EXPIRATION DATE: 09/07/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	22:41
INTERNAL STANDARD	VERIFIED	22:42
EXTERNAL STANDARD	.097	22:42
BLANK TEST	.000	22:43
EXTERNAL STANDARD	.099	22:43
BLANK TEST	.000	22:44
EXTERNAL STANDARD	.099	22:44
BLANK TEST	.000	22:45

N = 3
SIM. = .1
AVG. = .0983

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204131
01/14/14

ARREST TIME: 21:30
SUBJECT NAME:
TEST
DOB: 11/11/11 SEX: M
STATE/D.L.: MO/11111
ARRESTING OFFICER:
KIMBALL
OFFICER I.D.: 1149
TESTING OFFICER:
KIMBALL
OFFICER I.D.: 1149
PERMIT NUMBER: 220035
EXPIRATION DATE: 09/07/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:11
INTERNAL STANDARD	VERIFIED	22:11
RADIO INTERFERENCE		

Operator Signature

