



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 1:34 pm, Jun 04, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204137	NAME OF AGENCY MISSOURI STATE HIGHWAY PATROL	DATE OF INSPECTION 06/04/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 301 NORTH 2ND STREET, ST. CHARLES (ST. CHARLES COUNTY DOC)		TIME OF INSPECTION 9:20 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 06-04-14 09:55
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 13290 EXP. DATE 10/29/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.1 °C SIMULATOR SN G11079 EXP. DATE 10/16/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .103% TEST 2 .102% TEST 3 .103%

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	6	(.10-.14)	4	(.15-.19)	3	OVER .19	3
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE CPL. B. Seymour, #641	PRINT FULL NAME Corporal Barton M. Seymore, #641
------------------------------------	---

TYPE II PERMIT NUMBER/EXPIRATION DATE 220074 03/07/2016	TELEPHONE NUMBER (636) 300-2800
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, Missouri Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13290** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BARTON M SEYMORE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repair and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of section 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240074

EXPIRES 3/7/2016

MO 590-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAS-4 (R6-1)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator SEYMORE, BARTON
Permit No 240074
Date Issued 3/7/2014 Date Expires 3/7/2016

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204137
06/04/14
09:55

DIAGNOSTIC CHECK

COMPUTER: OKRY
PROGRAM (04-07-2009): OKRY
HEATERS
SAMPLE CHAMBER: 50C
FLOW DETECTOR: OKRY
PUMP
HIGH SPEED: OKRY
DETECTOR: OKRY
FILTERS: OKRY
PURTZ STANDARD: OKRY
CALIBRATION: OKRY

PRINTER TEST

!##\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~>

Operator Signature Mr. B. Simpson #641

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204137
06/04/14

TESTING OFFICER:

SEYMORE/B/M
OFFICER I.D.# 641
PERMIT NUMBER: 229074
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

SUPERVISOR MODE

BLANK TEST .000 10:17
INTERNAL STANDARD VERIFIED 10:17
EXTERNAL STANDARD .100 10:17
BLANK TEST .000 10:18
EXTERNAL STANDARD .100 10:18
BLANK TEST .000 10:19
EXTERNAL STANDARD .100 10:19
BLANK TEST .000 10:20

N = 3
SIM = .1
SWS = .1025

Operator Signature Mr. B. Simpson #641

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204137
05/04/14

ARREST TIME: 09:45
SUBJECT NAME:
RFI/TEST
DOB: 01/23/45 SEX: M
STATE/D.L.# MO/1234567890
ARRESTING OFFICER:
SEYMORE/B/M
OFFICER I.D.# 641
TESTING OFFICER:
SEYMORE/B/M
OFFICER I.D.# 641
PERMIT NUMBER: 220074
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

RADIO INTERFERENCE TEST

----- BREATH ANALYSIS -----

BLANK TEST .000 10:24
INTERNAL STANDARD VERIFIED 10:24
RADIO INTERFERENCE

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204137
06/04/14

ARREST TIME: 10:00
SUBJECT NAME:
BREATH/TEST
DOB: 12/25/67 SEX: M
STATE/D.L.# MO/0987654321
ARRESTING OFFICER:
SEYMORE/B/M
OFFICER I.D.# 641
TESTING OFFICER:
SEYMORE/B/M
OFFICER I.D.# 641
PERMIT NUMBER: 220074
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

RADIO INTERFERENCE TEST

----- BREATH ANALYSIS -----

BLANK TEST .000 10:39
INTERNAL STANDARD VERIFIED 10:39
SUBJECT SAMPLE .000 10:31
BLANK TEST .000 10:31

Operator Signature

Op. E. Seymour #641

Operator Signature

Op. E. Seymour #641