



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 3/5/14-CD

REPORT #6

REVIEWED

By Carol Day at 9:53 am, Mar 14, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days) or whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204141	NAME OF AGENCY MISSOURI STATE HIGHWAY PATROL	DATE OF INSPECTION 02-28-14
LOCATION OF INSTRUMENT (STREET AND CITY) 104 W. MAIN, WARRENTON, MO 63383		TIME OF INSPECTION 0941

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02-28-14 0941
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER REPO MARKETING LOT # 13001 EXP. DATE 03-07-15
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G6825 EXP. DATE 01/21/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .098	TEST 2 → .099	TEST 3 → .100
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 1	(.10-.14) 3	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME CPL. JEFFREY SCOTT GRAUE
TYPE II PERMIT NUMBER/EXPIRATION DATE 220149 06/26/2014	TELEPHONE NUMBER (636) 300-2800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REP CO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 200141
02/28/14

TESTING OFFICER:
GRAEBER, J/S
OFFICER I.D. # 154
PERMIT NUMBER: 200149
EXPIRATION DATE: 06/26/14
MISCELLANEOUS DATA:

----- SUPERVISOR MODE -----

BLANK TEST	.000	09:53
INTERNAL STANDARD	VERIFIED	09:53
EXTERNAL STANDARD	.098	09:54
BLANK TEST	.000	09:54
EXTERNAL STANDARD	.099	09:55
BLANK TEST	.000	09:55
EXTERNAL STANDARD	.100	09:56
BLANK TEST	.000	09:57

N = 3
STA. # 1
AVG. = .099

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 220148
02/28/14

ARREST TIME: 09:00

SUBJECT NAME:

DOB: 01/01/01 SEX: M

STATE D.L. #: MO-122456

ARRESTING OFFICER:

OFFICER I.D. #: X

TESTING OFFICER:

GRADE: J/S

OFFICER I.D. #: 154

PERMIT NUMBER: 220148


EXPIRATION DATE: 06/26/14

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	09:59
INTERNAL STANDARD	VERIFIED	09:59
RADIO INTERFERENCE		

Operator Signature



State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



J S GRAUE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1988.

Date 06/26/2012


Number 220149

Expires 06/26/2014

MO 880-0771 (7-88)



Director of State Public Health Laboratory



Director, Department of Health

Lab. 4 (17-88)