



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**  
 REPORT #6  
 By Carol Day at 8:58 am, Apr 01, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days) or at the time of the regular monthly preventive maintenance check (not to exceed 30 days) or at the time of the regular monthly preventive maintenance check (not to exceed 30 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204144	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 03/31/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Cooper County Jail, Boonville		TIME OF INSPECTION 2:09 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 03/31/2014, 14:09
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>RepCo Marketing Inc.</u> LOT # <u>13001</u> EXP. DATE <u>03/07/2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C SIMULATOR SN <u>G6784</u> EXP. DATE <u>08/20/2014</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ➔ .103%	TEST 2 ➔ .103%	TEST 3 ➔ .103%
----------------	----------------	----------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument operated within DHSS standards.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME Corporal Ryan S. Thompson
TYPE II PERMIT NUMBER/EXPIRATION DATE 230337 12/23/2015	TELEPHONE NUMBER (573) 751-1000

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

Face This Side Down - This Edge In First

**BAC DataMaster**  
**Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204144  
03/31/14  
14:09

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 49c  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopq  
rstuvwxyz{|}~

Operator Signature \_\_\_\_\_



Face This Side Down - This Edge In First

**BAC DataMaster**  
**Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204144  
03/31/14

TESTING OFFICER:  
THOMPSON/RYAN/S  
OFFICER I.D.: 672  
PERMIT NUMBER: 230337  
EXPIRATION DATE: 12/23/15  
MISCELLANEOUS DATA:  
D

--- SUPERVISOR MODE ---

BLANK TEST	.000	14:13
INTERNAL STANDARD	VERIFIED	14:13
EXTERNAL STANDARD	.103	14:13
BLANK TEST	.000	14:14
EXTERNAL STANDARD	.103	14:14
BLANK TEST	.000	14:15
EXTERNAL STANDARD	.103	14:15
BLANK TEST	.000	14:16

N = 3  
SIM. = .1  
AVG. = .103

Operator Signature \_\_\_\_\_



Face This Side Down - This Edge In First

**BAC DataMaster**  
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204144  
03/31/14

ARREST TIME: 10:00  
SUBJECT NAME:  
MMMM  
DOB: 11/11/11      SEX: M  
STATE/D.L.: MO/1111  
ARRESTING OFFICER:  
000  
OFFICER I.D.: 111  
TESTING OFFICER:  
THOMPSON/RYAN/S  
OFFICER I.D.: 672  
PERMIT NUMBER: 230337  
EXPIRATION DATE: 12/23/15  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**RYAN S THOMPSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/23/2013  
 NUMBER 230337  
 EXPIRES 12/23/2015

  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
*Gail Vosterly*  
 ,acting director  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator THOMPSON, RYAN  
 Permit No 230337  
 Date Issued 12/23/2013 Date Expires 12/23/2015

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13001**

**EXPIRATION DATE: March 7, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.