MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 8:55 am, Dec 08, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
204145	MSHP	11/30/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
Nodaway County Sheriff's Office, Maryville, Missouri	8:37 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 11/30/2014, 0837

☑ COMPUTER ☑ DETECTOR

☑ PROGRAM ☑ FILTERS

☑ HEATERS SAMPLE CHAMBER 50 °C ☑ QUARTZ STANDARD

☑ FLOW DETECTOR ☑ CALIBRATION

☑ PUMP HIGH SPEED ☑ PRINTER

☑ INDICATOR LIGHTS

☑ SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13280 EXP. DATE 10/16/2015

☑ SIMULATOR TEMP (34°C ± 0.2°C) 34.01 °C SIMULATOR SN G8005 EXP. DATE 07/14/2015

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛	TEST 2 ☛	TEST 3 ☛
.100	100	.100

☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
	0	0	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

MEETS DOH Specs. Lo #13280, Bot. #24

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
D. R. Reuter	D. R. Reuter

TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240106 03/11/2016	(816) 387-2345

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFIED ALCOHOL REFERENCE SOLUTION FOR SIMULATOR

13280	10/16/13	10/16/15
LOT NO.	MFG. DATE	EXP. DATE

275 Gal.	500 ML	
LOT VOL.	BOT. VOL.	BOT. NO.

When this reference solution is used with a breath simulator certified by Guth Laboratories, a properly operating instrument will read 0.10

For additional information contact:

Guth Laboratories, Inc.
590 North 67th Street, Harrisburg, PA 17111
Toll Free 800-233-2338

Rev. 4/02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204145
11/30/14

TESTING OFFICER:
REUTER/DALE/R
OFFICER I.D.: 622
PERMIT NUMBER: 240106
EXPIRATION DATE: 03/11/16
MISCELLANEOUS DATA:
TEST

--- SUPERVISOR MODE ---

BLANK TEST	.000	09:06
INTERNAL STANDARD	VERIFIED	09:06
EXTERNAL STANDARD	.100	09:07
BLANK TEST	.000	09:07
EXTERNAL STANDARD	.100	09:08
BLANK TEST	.000	09:09
EXTERNAL STANDARD	.100	09:09
BLANK TEST	.000	09:10

N = 3
SIM. = .1
AVG. = .1

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 000145
11/30/14
08:51

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	50c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204145
11/30/14

ARREST TIME: 00:00
SUBJECT NAME:
SMITH/JOE/P
DOB: 01/02/87 SEX: M
STATE/D.L.: MO/123456789O
ARRESTING OFFICER:
REUTER/DALE/R
OFFICER I.D.: 622
TESTING OFFICER:
REUTER/DALE/R
OFFICER I.D.: 622
PERMIT NUMBER: 240106
EXPIRATION DATE: 03/11/16
MISCELLANEOUS DATA:
RFI
RFI

--- BREATH ANALYSIS ---

BLANK TEST .000 08:56
INTERNAL STANDARD VERIFIED 08:56
RADIO INTERFERENCE

Operator Signature

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

DALE R REUTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 04/02/2013

NUMBER 230058

EXPIRES 04/02/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)