



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 3/5/14-CD

REVIEWED
 By Carol Day at 3:52 pm, Mar 20, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|--|---|----------------------------------|
| DATAMASTER SN 204148 | NAME OF AGENCY Missouri State Highway Patrol | DATE OF INSPECTION 02/28/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 110 W. Elm St. Room 70 Ozark | | TIME OF INSPECTION 2143 |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

| | |
|---|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) | DATE AND TIME (from printout) <u>02/28/14 2143</u> |
| <input checked="" type="checkbox"/> COMPUTER | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTERS |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR | <input checked="" type="checkbox"/> CALIBRATION |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED | <input checked="" type="checkbox"/> PRINTER |
| <input checked="" type="checkbox"/> INDICATOR LIGHTS | |
| <input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>RepCo Marketing Inc.</u> LOT # <u>13001</u> EXP. DATE <u>03/07/2015</u> | |
| <input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.2</u> °C SIMULATOR SN <u>G11046</u> EXP. DATE <u>03/26/2014</u> | |
| <input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) | |
| Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) | |
| <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE | |
| <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE | |
| <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE | |

| | | |
|--|--|--|
| TEST 1 <input checked="" type="checkbox"/> <u>.095 %</u> | TEST 2 <input checked="" type="checkbox"/> <u>.097 %</u> | TEST 3 <input checked="" type="checkbox"/> <u>.098 %</u> |
|--|--|--|

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | |
|----------|-----------|-------------|-------------|-------------|------------|---|
| REFUSALS | 0 (0-.04) | 0 (.05-.09) | 0 (.10-.14) | 0 (.15-.19) | 0 OVER .19 | 1 |
|----------|-----------|-------------|-------------|-------------|------------|---|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Alcohol Vapor Concentration .100% Bottle #318 Lot #13001 RepCo Marketing Inc.

| | |
|--|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE <i>Michael W. Lee</i> | PRINT FULL NAME Michael W. Lee |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 220217 09/04/2014 | TELEPHONE NUMBER (417) 895-6868 |

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.03 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



MICHAEL W LEE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/04/2012

Number 220217

Expires 09/04/2014

Director of State Public Health Laboratory

Director, Department of Health

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATA MASTER SERIAL NUMBER 204103
60/08/14
21145

--- DEPARTMENTAL CONTROL ---

COMPUTER: UNIT

PROGRAM: UNIT

HEADERS: UNIT

CRIMINAL: UNIT

FLOW: UNIT

PUMP: UNIT

HTOH: UNIT

DIFFERENCE: UNIT

FLOW: UNIT

QUANTITY: UNIT

CHILDREN: UNIT

PRINTING UNIT

MISSOURI STATE HIGHWAY PATROL
BAC DATA MASTER SERIAL NUMBER 204103
60/08/14
21145

Operator Signature M. W. LEE

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATA MASTER SERIAL NUMBER 204103
60/08/14

TESTING OFFICER:

LEE, M. W.

OFFICER ID: 1012

PERMIT NUMBER: 1012

OPERATION DATE: 60/08/14

MISCELLANEOUS: UNIT

--- SUPERVISOR CONTROL ---

| | | |
|-------------------|-------|-------|
| BLANK TEST | 1.000 | 21150 |
| INTERNAL STANDARD | 1.000 | 21150 |
| EXTERNAL STANDARD | 1.000 | 21154 |
| BLANK TEST | 1.000 | 21155 |
| INTERNAL STANDARD | 1.000 | 21155 |
| BLANK TEST | 1.000 | 21156 |
| EXTERNAL STANDARD | 1.000 | 21156 |
| BLANK TEST | 1.000 | 21157 |

UNIT

UNIT

UNIT

Operator Signature M. W. LEE

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204193
02/28/14

ARREST TIME: 20:30
SUBJECT NAME:
TEST#(S):
DOB: 01/01/90 SEX: M
STATE/CLL# NO: 123456
ARRESTING OFFICER:
LEE/M/W
OFFICER I.D.#: 842
TESTING OFFICER:
LEE/M/W
OFFICER I.D.#: 842
PERMIT NUMBER: 200212
EXPIRATION DATE: 09-09-14
MISCELLANEOUS DATA:

BREATH ANALYSIS

| | | |
|--------------------|----------|-------|
| BLANK TEST | 100% | 02:00 |
| INTERNAL STANDARD | VERIFIED | 02:00 |
| ADULT INTERFERENCE | | |

Operator Signature M. W. LEE