

DOH



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
DATAMASTER MAINTENANCE REPORT

RECEIVED 7/14/14-CD

REVIEWED  
By Carol Day at 9:47 am, Jul 22, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204150 Inv.#127265	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 07/06/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 226 Main St., Forsyth (Taney County Jail)		TIME OF INSPECTION 9:43 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07/06/2014 0943
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 13280 EXP. DATE 10/16/2015

<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN G11057 EXP. DATE 01/16/2015
--

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .097%	TEST 2 • .097%	TEST 3 • .097%
----------------	----------------	----------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	1	(.05-.09)	0	(.10-.14)	1	(.15-.19)	7	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Operates within the guidelines established by the Missouri Department of Health

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME Tpr. Terry A. Bible
TYPE II PERMIT NUMBER/EXPIRATION DATE 240044 03/07/2016	TELEPHONE NUMBER (417) 895-6868

RETURN COMPLETED REPORT TO THE:  
Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-6470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**TERRY A BIBLE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240044

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator **BIBLE, TERRY**  
 Permit No **240044**  
 Date Issued **3/7/2014**    Date Expires **3/7/2016**

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204150  
07/06/14  
09:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 50c  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~

Operator Signature

*J. B. L.* #264

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204150  
07/06/14

TESTING OFFICER:  
BIBBLE/TERRY/A  
OFFICER I.D.#: 264  
PERMIT NUMBER: 24004  
EXPIRATION DATE: 03/07/16  
MISCELLANEOUS DATA:  
TEST .10

--- SUPERVISOR MODE ---

BLANK TEST	.000	09:46
INTERNAL STANDARD	VERIFIED	09:46
EXTERNAL STANDARD	.097	09:46
BLANK TEST	.000	09:47
EXTERNAL STANDARD	.097	09:47
BLANK TEST	.000	09:48
EXTERNAL STANDARD	.097	09:48
BLANK TEST	.000	09:49

n = 3  
S.D. = .1  
AVG. = .097

Operator Signature

*J. B. L.* #264

Face This Side Down - This Edge In First

**BAC DataMaster**  
**Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204150  
07/06/14

ARREST TIME: 01:11  
SUBJECT NAME:  
BIBLE/T/A  
DOB: 01/11/11 SEX: M  
STATE/D.L.: MO/0123456789  
ARRESTING OFFICER:  
BIBLE/T/A  
OFFICER I.D.: 264  
TESTING OFFICER:  
SAME  
OFFICER I.D.: SAME  
PERMIT NUMBER: 24004  
EXPIRATION DATE: 03/07/16  
MISCELLANEOUS DATA:  
TEST RFI

--- BREATH ANALYSIS ---

BLANK TEST	.000	09:52
INTERNAL STANDARD	VERIFIED	09:52
RADIO INTERFERENCE		

Operator Signature

*Jas Bible #264*