



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 2/19/14-CD **REPORT #6**

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days) or whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 10:02 am, Mar 14, 2014

DATAMASTER SN 204154	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/13/2014
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Moniteau County Jail, 210 North Street, California, MO 65018	TIME OF INSPECTION 11:54 am
--	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 02-13-2014, 1154
- COMPUTER DETECTOR
- PROGRAM FILTERS
- HEATERS SAMPLE CHAMBER 49 °C QUARTZ STANDARD
- FLOW DETECTOR CALIBRATION
- PUMP HIGH SPEED PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP ($34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$) 33.99 °C SIMULATOR SN G6759 EXP. DATE 02/07/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within $\pm 5\%$ of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 \rightarrow 0.100	TEST 2 \rightarrow 0.100	TEST 3 \rightarrow 0.100
----------------------------	----------------------------	----------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	2	(.10-.14)	1	(.15-.19)	2	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument meets Department of Health Regulations.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Brian J. Geier
---------------	-----------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230324 12/23/2015	TELEPHONE NUMBER (573) 751-1000
--	------------------------------------

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901**

REPco MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5400

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013

The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204154
02/13/14
11:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49C
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~

Operator Signature *Tracy L. #599*

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204154
02/13/14

ARREST TIME: 11:00

SUBJECT NAME:

TEST

DOB: 01/01/01 SEX: M

STATE/D.L.: MO/TEST

ARRESTING OFFICER:

TEST

OFFICER I.D.: TEST

TESTING OFFICER:

GEIER/BRIAN/J

OFFICER I.D.: 599

PERMIT NUMBER: 230324

EXPIRATION DATE: 12/23/15

MISCELLANEOUS DATA:

RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 12:03
INTERNAL STANDARD VERIFIED 12:03
RADIO INTERFERENCE

Operator Signature *Tracy L. #599*



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

BRIAN J GEIER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit Issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/23/2013

NUMBER 230324

EXPIRES 12/23/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (06-10)

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204154
02/13/14

TESTING OFFICER:
GEIER/BRIAN/J
OFFICER I.D.: 599
PERMIT NUMBER: 230324
EXPIRATION DATE: 12/23/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	11:57
INTERNAL STANDARD	VERIFIED	11:57
EXTERNAL STANDARD	.100	11:58
BLANK TEST	.000	11:58
EXTERNAL STANDARD	.100	11:59
BLANK TEST	.000	11:59
EXTERNAL STANDARD	.100	12:00
BLANK TEST	.000	12:00

N = 3

SIM. = .1

AVG. = .1

Operator Signature