



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 2/19/14-CD

REPORT #6

REVIEWED

By Carol Day at 10:02 am, Mar 14, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 1 month).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204158	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/13/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Morgan County Jail, 211 E. Newton, Versailles, MO 65084		TIME OF INSPECTION 1:13 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>02-13-2014, 1313</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>RepCo Marketing</u>	LOT # <u>13001</u> EXP. DATE <u>03/07/2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>33.99</u> °C	SIMULATOR SN <u>G6759</u> EXP. DATE <u>02/07/2015</u>
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	

TEST 1 <input checked="" type="checkbox"/> 0.101	TEST 2 <input checked="" type="checkbox"/> 0.102	TEST 3 <input checked="" type="checkbox"/> 0.102
--	--	--

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(.0-.04)	2	(.05-.09)	0	(.10-.14)	1	(.15-.19)	4	OVER .19	0
----------	---	----------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument meets Department of Health Regulations.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Brian J. Geier
TYPE II PERMIT NUMBER/EXPIRATION DATE 230324 12/23/2015	TELEPHONE NUMBER (573) 751-1000

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REPCO MARKETING INC.2101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-878-5480**CERTIFICATE OF ANALYSIS**

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204158
02/13/14

TESTING OFFICER:
GEYER/BRIAN/J
OFFICER I.D.: 599
PERMIT NUMBER: 230324
EXPIRATION DATE: 12/23/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	13:41
INTERNAL STANDARD	VERIFIED	13:41
EXTERNAL STANDARD	.101	13:41
BLANK TEST	.000	13:42
EXTERNAL STANDARD	.102	13:42
BLANK TEST	.000	13:43
EXTERNAL STANDARD	.102	13:43
BLANK TEST	.000	13:44

N = 3
SIM. = .1
AVG. = .1016

Operator Signature: *[Signature]* 78599

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204158
02/13/14
13:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 50C

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! "#%&'()*+,-./0123456789:;<=>?@ABCDEF G

H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o

p q r s t u v w x y z { | } ~

Operator Signature: *[Signature]* 78599



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRIAN J GEIER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/23/2013

NUMBER 230324

EXPIRES 12/23/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (PB-10)

MO 580-0771 (6-10)

BAC DataMaster
Evidence Ticket

LAB-4 (PB-10) MISSOURI STATE HIGHWAY PATROL

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204156
02/18/14

ARREST TIME: 12:00
SUBJECT NAME:
TEST

DOB: 01/01/01 SEX: M
STATE/D.L.: MP/TEST
ARRESTING OFFICER:
TEST

OFFICER I.D.: TEST
TESTING OFFICER:
GEIER/BRIAN/J
OFFICER I.D.: 599

PERMIT NUMBER: 230324
EXPIRATION DATE: 12/23/15
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST
INTERNAL STANDARD
RADIO INTERFERENCE
VERIFIED
13:47
13:47

Operator Signature