



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 3:34 pm, Jan 06, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204160	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 01/03/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Clay County Detention Center - 14 South Water, Liberty, Missouri 64068		TIME OF INSPECTION 5:53 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01/03/2014 17:53</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Repco Marketing INC.</u> LOT # <u>12002</u> EXP. DATE <u>08/29/2014</u>

<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C SIMULATOR SN <u>G11069</u> EXP. DATE <u>10/03/2014</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <input checked="" type="checkbox"/> .100 %	TEST 2 <input checked="" type="checkbox"/> .101 %	TEST 3 <input checked="" type="checkbox"/> .101 %
---	---	---

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	3	(0-.04)	0	(.05-.09)	1	(.10-.14)	5	(.15-.19)	0	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Ryan J. Wilhoit
TYPE II PERMIT NUMBER/EXPIRATION DATE 230313 12/11/2015	TELEPHONE NUMBER (816) 622-0800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MICHIGAN STATE HIGHWAY PATROL
ONE DETROITER SERIAL NUMBER 204109
01/03/14

ISSUING OFFICER:
DALLAN R/J

OFFICER I.D.#: 1129
SERIAL NUMBER: 230313
EXPIRATION DATE: 12/11/15
MISCELLANEOUS DATA:
100% SOLUTION

--- SUPERVISOR MODE ---

STAMP TEST	.000	18:21
STAMP TEST	VERIFIED	18:21
STAMP TEST	.100	18:22
STAMP TEST	.000	18:22
STAMP TEST	.101	18:23
STAMP TEST	.000	18:23
STAMP TEST	.101	18:24
STAMP TEST	.000	18:24

100% SOLUTION

Operator Signature

PR. R.J. DALLAN #1129

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MICHIGAN STATE HIGHWAY PATROL
ONE DETROITER SERIAL NUMBER 204109
01/03/14
17:53

--- DIAGNOSTIC CHECK ---

CONTROLLER: OKAY
FUNCTION (94-07-2009): OKAY
MEMORY: OKAY
SERIAL NUMBER: 500
STAMP TEST: OKAY

STAMP TEST: OKAY
STAMP TEST: OKAY
STAMP TEST: OKAY
STAMP TEST: OKAY
STAMP TEST: OKAY
STAMP TEST: OKAY
STAMP TEST: OKAY
STAMP TEST: OKAY
STAMP TEST: OKAY
STAMP TEST: OKAY

100% SOLUTION

100% SOLUTION

FRANKE TEST
SERIAL NUMBER: 0123456789
MISCELLANEOUS DATA: abcdefghijklmnop
01/03/14

Operator Signature

PR. R.J. DALLAN #1129

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
PCC IDENTIFIER SERIAL NUMBER: 204160
01/03/14

ARREST TIME: 17:45
SUBJECT NAME:
RILEY, J.
DOB: 02/03/82 SEX: M
MOTOR NO: 123456789
CHASING OFFICER:
SHERIFF, R. J.
OFFICER I.D.#: 1129
TESTING OFFICER:
SHERIFF, R. J.
OFFICER I.D.#: 1129
EVIDENCE NUMBER: 204160
EVIDENCE DATE: 12/11/13
MISCELLANEOUS NOTES:
(NONE)

BREATH ANALYSIS

ALCOHOL	0.000	13:29
DRUGS	NEGATIVE	13:29
TOXIC		

Operator Signature: PR. R. J. [Signature] #1129