MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 1:07 pm, Feb 07, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
204160	Missouri State Highway Patrol	02/03/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		**TIME OF INSPECTION**
Clay County Detention Center - 14 South Water, Liberty, Missouri 64068		9:28 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 02/03/2014 09:28
- ☑ COMPUTER — ☑ DETECTOR
- ☑ PROGRAM — ☑ FILTERS
- ☑ HEATERS SAMPLE CHAMBER 50 °C — ☑ QUARTZ STANDARD
- ☑ FLOW DETECTOR — ☑ CALIBRATION
- ☑ PUMP HIGH SPEED — ☑ PRINTER
- ☑ INDICATOR LIGHTS
- ☑ SIMULATOR SOLUTION SUPPLIER Repco Marketing Inc. LOT # 12002 EXP. DATE 08/29/2014
- ☑ SIMULATOR TEMP (34°C ± 0.2°C) 34.2 °C SIMULATOR SN G11069 EXP. DATE 10/03/2014
- ☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.100 %	.100 %	.101 %

- ☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
4	0	4	6	3	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸ R. Wilhoit #1129	Ryan J. Wilhoit
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
230313 12/11/2015	(816) 622-0800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 12002
EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **12002** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1209** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **August 30, 2012** The expiration date for this lot number is **August 29, 2014** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204160
02/03/14

TESTING OFFICER:
WILHOIT/R/J
OFFICER I.D.: 1129
PERMIT NUMBER: 230313
EXPIRATION DATE: 12/11/15
MISCELLANEOUS DATA:
.100% SOLUTION

--- SUPERVISOR MODE ---

BLANK TEST	.000	09:43
INTERNAL STANDARD	VERIFIED	09:43
EXTERNAL STANDARD	.100	09:44
BLANK TEST	.000	09:44
EXTERNAL STANDARD	.100	09:45
BLANK TEST	.000	09:45
EXTERNAL STANDARD	.101	09:46
BLANK TEST	.000	09:46

N = 2
AVG. = .1
AVG. = .1003

Operator Signature IPR. R. J. WILHOIT #1129

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204160
02/03/14
09:28

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS:	
SAMPLE CHAMBER:	50C
FLOW DETECTOR:	OKAY
PUMP:	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
AUXILIARY STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature IPR. R. J. WILHOIT #1129

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204160
02/03/14

ARREST TIME: 09:00
SUBJECT NAME:
RFI/TEST
DOB: 02/03/82 SEX: N
STATE/D.L.: MO 1234567890
ARRESTING OFFICER:
WILHOIT/R/J
OFFICER I.D.: 1129
TESTING OFFICER:
WILHOIT/R/J
OFFICER I.D.: 1129
PERMIT NUMBER: 230313
EXPIRATION DATE: 12/11/15
MISCELLANEOUS DATA:
RFI TEST

---- BREATH ANALYSIS ----

BLANK TEST	.000	09:51
INTERNAL STANDARD	VERIFIED	09:51
RADIO INTERFERENCE		

Operator Signature PRT R. J. WILHOIT #1129