



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 3/6/14-CD REPORT #6

REVIEWED
 By Carol Day at 3:23 pm, Apr 01, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed in service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204163	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Carroll County Sheriff's Department, 106 S. Folger St., Carrollton, MO 64633		TIME OF INSPECTION 1120

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02/28/14, 11:28
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marketing, Inc. LOT # 13001 Bot. 0738 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.2 °C SIMULATOR SN G11067 EXP. DATE 10-4-2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 - .103	TEST 2 - .104	TEST 3 - .104
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 1	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE <i>Cpl. Neil K. Johnson #1112</i>	PRINT FULL NAME Cpl. Neil K. Johnson #1112
--	---

TYPE II PERMIT NUMBER/EXPIRATION DATE 220181 08/03/2014	TELEPHONE NUMBER (816) 622-0800
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
02/28/14
11:28

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 50c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

Operator Signature

2208-02

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
02/28/14

TESTING OFFICER:
JOHNSON/NEIL/K
OFFICER I.D.: 1112
PERMIT NUMBER: 220181
EXPIRATION DATE: 08/03/14
MISCELLANEOUS DATA:
REPCO .10, LOT 13001
BOT. 0738, EXP. 3-7-15

--- SUPERVISOR MODE ---

BLANK TEST	.000	11:39
INTERNAL STANDARD	VERIFIED	11:40
EXTERNAL STANDARD	.103	11:40
BLANK TEST	.000	11:41
EXTERNAL STANDARD	.104	11:41
BLANK TEST	.000	11:42
EXTERNAL STANDARD	.104	11:42
BLANK TEST	.000	11:43

N = 3
SIM. = .1
AVG. = .1036

Operator Signature

2208-02

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
02/28/14

ARREST TIME: 11:20
SUBJECT NAME:
ME
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/0987654321
ARRESTING OFFICER:
ME
OFFICER I.D.: ME
TESTING OFFICER:
JOHNSON/NEIL/K
OFFICER I.D.: 1112
PERMIT NUMBER: 220181
EXPIRATION DATE: 08/03/14
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	11:47
INTERNAL STANDARD	VERIFIED	11:47
RADIO INTERFERENCE		

Operator Signature



2208-02

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
02/28/14

ARREST TIME: 11:20
SUBJECT NAME:
ME
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/0987654321
ARRESTING OFFICER:
ME
OFFICER I.D.: ME
TESTING OFFICER:
JOHNSON/NEIL/K
OFFICER I.D.: 1112
PERMIT NUMBER: 220181
EXPIRATION DATE: 08/03/14
MISCELLANEOUS DATA:
SELF TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	11:50
INTERNAL STANDARD	VERIFIED	11:50
SUBJECT SAMPLE	.000	11:51
BLANK TEST	.000	11:51

Operator Signature



2208-02