MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 12:34 pm, Apr 07, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
204163	Missouri State Highway Patrol	04/04/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
Carroll County Sheriff's Department, 106 S. Folger St., Carrollton, MO 64633	6:44 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 04-04-2014 at 0644 hrs
- ☑ COMPUTER — ☑ DETECTOR
- ☑ PROGRAM — ☑ FILTERS
- ☑ HEATERS SAMPLE CHAMBER 50 °C — ☑ QUARTZ STANDARD
- ☑ FLOW DETECTOR — ☑ CALIBRATION
- ☑ PUMP HIGH SPEED — ☑ PRINTER
- ☑ INDICATOR LIGHTS
- ☑ SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 13001 EXP. DATE 03/07/2015
- ☑ SIMULATOR TEMP (34°C ± 0.2°C) 34.1 °C SIMULATOR SN G11041 EXP. DATE 09/09/2014
- ☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

  - ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
  - ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
  - ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛ .102 BAC	TEST 2 ☛ .102 BAC	TEST 3 ☛ .102 BAC

- ☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	1	0	0	1

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Operating within MODHSS specifications
Solution bottle # 0597

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸ TPR [signature]	TPR C.D. Moeller #662
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
240064 03/07/2016	(816) 622-0800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscrimatory basis

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

CLAY D MOELLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240064

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOELLER, CLAY
Permit No 240064
Date Issued 3/7/2014 **Date Expires** 3/7/2016

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **13001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1215** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **March 8, 2013** The expiration date for this lot number is **March 7, 2015** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
04/04/14
06:44

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	50c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

```
 !"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~
```

Operator Signature TPR

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
04/04/14

TESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
PERMIT NUMBER: 240064
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
INV# 127285

--- SUPERVISOR MODE ---

BLANK TEST	.000	06:46
INTERNAL STANDARD	VERIFIED	06:46
EXTERNAL STANDARD	.102	06:47
BLANK TEST	.000	06:48
EXTERNAL STANDARD	.102	06:48
BLANK TEST	.000	06:49
EXTERNAL STANDARD	.102	06:49
BLANK TEST	.000	06:50

N = 3
SIM. = .1
AVG. = .102

Operator Signature TPR

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
04/04/14

ARREST TIME: 00:01
SUBJECT NAME:
RFI/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/X
ARRESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
TESTING OFFICER:
MOELLER/C/
OFFICER I.D.: 0662
PERMIT NUMBER: 240064
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
INV# 127285

--- BREATH ANALYSIS ---

BLANK TEST	.000	06:52
INTERNAL STANDARD	VERIFIED	06:52
RADIO INTERFERENCE		

Operator Signature TPR CJM

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
04/04/14

ARREST TIME: 00:01
SUBJECT NAME:
BLANK/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/X
ARRESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
TESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
PERMIT NUMBER: 240064
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
INV# 127285

--- BREATH ANALYSIS ---

BLANK TEST	.000	06:54
INTERNAL STANDARD	VERIFIED	06:54
SUBJECT SAMPLE	.000	06:55
BLANK TEST .	.000	06:55

Operator Signature TPR CJM

2208-02