MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 11:24 am, Dec 08, 2014
REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
204163	Missouri State Highway Patrol	12/01/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
106 S. Folger St., Carrollton, Missouri 64633	3:13 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 12/01/2014 at 1513 hrs

☑ COMPUTER	☑ DETECTOR
☑ PROGRAM	☑ FILTERS
☑ HEATERS SAMPLE CHAMBER 50 °C	☑ QUARTZ STANDARD
☑ FLOW DETECTOR	☑ CALIBRATION
☑ PUMP HIGH SPEED	☑ PRINTER

☑ INDICATOR LIGHTS

☑ SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 13290 EXP. DATE 10/29/2015

☑ SIMULATOR TEMP (34°C ± 0.2°C) 33.99 °C SIMULATOR SN MP2125 EXP. DATE 06/03/2015

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛	TEST 2 ☛	TEST 3 ☛
.097	.098	.097

☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	0	1	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Operating within MODHSS specifications
Solution bottle # 767, Guth 12V500 Simulator

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸ Tpr. J.T.	Tpr. John T. Cunningham
TYPE II PERMIT NUMBER/EXPIRATION DATE: 240386 11/13/2016	TELEPHONE NUMBER: (816) 622-0800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscrimatory basis

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13290** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FN122211-02** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

```
MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
12/01/14
15:13

--- DIAGNOSTIC CHECK ---

COMPUTER:                    OKAY

PROGRAM (04-07-2009):        OKAY

HEATERS
SAMPLE CHAMBER:              50c

FLOW DETECTOR:               OKAY

PUMP
HIGH SPEED:                  OKAY

DETECTOR:                    OKAY

FILTERS:                     OKAY

QUARTZ STANDARD:             OKAY

CALIBRATION:                 OKAY

                PRINTER TEST
 !"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~
```

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

```
MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
12/01/14

TESTING OFFICER:
  CUNNINGHAM/JOHN/T
OFFICER I.D.: 561
PERMIT NUMBER: 240386
EXPIRATION DATE: 11/13/16
MISCELLANEOUS DATA:

        --- SUPERVISOR MODE ---

BLANK TEST                .000          15:20
INTERNAL STANDARD         VERIFIED      15:20
EXTERNAL STANDARD         .097          15:23
BLANK TEST                .000          15:24
EXTERNAL STANDARD         .098          15:25
BLANK TEST                .000          15:25
EXTERNAL STANDARD         .097          15:26
BLANK TEST                .000          15:26

N = 3
SIM. = .:
AVG. = .0973
```

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

```
MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204163
12/01/14

ARREST TIME: 15:00
SUBJECT NAME:
  SMITH/JOHN/M
DOB: 01/01/90          SEX: M
STATE/D.L.#: MO/XXXXXX
ARRESTING OFFICER:
  CUNNINGHAM/JOHN/T
OFFICER I.D.: 561
TESTING OFFICER:
  CUNNINGHAM/JOHN/T
OFFICER I.D.: 561
PERMIT #NUMBER: 240386
EXPIRATION DATE: 11/13/16
MISCELLANEOUS DATA:

        --- BREATH ANALYSIS ---

BLANK TEST                .000          15:20
INTERNAL STANDARD         VERIFIED      15:20
RADIO INTERFERENCE
```

Operator Signature

CMSU 2208-02

Printed on recycled paper with agri-based inks

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

JOHN T CUNNINGHAM

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/13/2014

NUMBER 240386

EXPIRES 11/13/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CUNNINGHAM, JOHN
Permit No 240386
Date Issued 11/13/2014 **Date Expires** 11/13/2016