



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 1:01 pm, Jun 19, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204168	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 06/10/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Saline County Jail, 1915 W. Arrow St., Marshall, MO 65340		TIME OF INSPECTION 5:03 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>06-10-2014 at 1703 hrs</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.1 °C SIMULATOR SN G11041 EXP. DATE 09/09/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <u>.103</u> BAC	TEST 2 <u>.103</u> BAC	TEST 3 <u>.103</u> BAC
------------------------	------------------------	------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Operating within MODHSS specifications
 Solution bottle # 0596

INSPECTING OFFICER

SIGNATURE <i>TPR C.D. Moeller</i>	PRINT FULL NAME TPR C.D. Moeller #662
TYPE II PERMIT NUMBER/EXPIRATION DATE 240064 03/07/2016	TELEPHONE NUMBER (816) 622-0800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CLAY D MOELLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240064

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator MOELLER, CLAY
Permit No 240064
Date Issued 3/7/2014 **Date Expires** 3/7/2016

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204168
06/10/14
17:03

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 50c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!##%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopq
rstuvwxyz{|}~

Operator Signature TPR CA M...

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204168
06/10/14

TESTING OFFICER:

MOELLER/C/D

OFFICER I.D.: 0662

PERMIT NUMBER: 240064

EXPIRATION DATE: 03/07/16

MISCELLANEOUS DATA:

INV# 127279

--- SUPERVISOR MODE ---

BLANK TEST	.000	17:25
INTERNAL STANDARD	VERIFIED	17:25
EXTERNAL STANDARD	.103	17:25
BLANK TEST	.000	17:26
EXTERNAL STANDARD	.103	17:26
BLANK TEST	.000	17:27
EXTERNAL STANDARD	.104	17:27
BLANK TEST	.000	17:28

N = 3
SIM. = .1
AVG. = .1033

Operator Signature TPR CA M...

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204168
06/10/14

ARREST TIME: 00:01
SUBJECT NAME:
RFI/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/X
ARRESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
TESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
PERMIT NUMBER: 240064
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
INV# 127279

--- BREATH ANALYSIS ---

BLANK TEST .000 17:30
INTERNAL STANDARD VERIFIED 17:30
RADIO INTERFERENCE

Operator Signature TRE CO MUE

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204168
06/10/14

ARREST TIME: 00:01
SUBJECT NAME:
BLANK/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/X
ARRESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
TESTING OFFICER:
MOELLER/C/D
OFFICER I.D.: 0662
PERMIT NUMBER: 240064
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
INV# 127279

--- BREATH ANALYSIS ---

BLANK TEST .000 17:32
INTERNAL STANDARD VERIFIED 17:32
SUBJECT SAMPLE .000 17:33
BLANK TEST .000 17:33

Operator Signature TRE CO MUE