



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 2/18/14 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 1:39 pm, Apr 01, 2014

DATAMASTER SN 204177	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/10/2014
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Troop D Satellite, Carthage, Missouri	TIME OF INSPECTION 2:38 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 02-10-14 2:38pm

COMPUTER DETECTOR

PROGRAM FILTERS

HEATERS SAMPLE CHAMBER 49 °C QUARTZ STANDARD

FLOW DETECTOR CALIBRATION

PUMP HIGH SPEED PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G11047 EXP. DATE 05/02/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .103	TEST 2 .103	TEST 3 .105
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	3	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This unit is operating within the rules and regulations of the Missouri Department of Health and Senior Services.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Jeffery L. Prewitt
---------------	---------------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 220269 09/13/2014	TELEPHONE NUMBER (417) 895-6868
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
02/10/14

TESTING OFFICER:
PREMITT
OFFICER I.D.: 981
PERMIT NUMBER: 220269
EXPIRATION DATE: 09/13/14
MISCELLANEOUS DATA:
TEST

--- SUPERVISOR MODE ---

BLANK TEST	.000	VERIFIED	14:51
INTERNAL STANDARD	.103		14:51
EXTERNAL STANDARD	.000		14:52
BLANK TEST	.123		14:53
EXTERNAL STANDARD	.000		14:53
BLANK TEST	.125		14:54
EXTERNAL STANDARD	.000		14:54
BLANK TEST			14:55

N = 3
Stm. = .1
RMS. = .1036

Operator Signature *[Signature]*

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
02/10/14

ARREST TIME: 12:00
SUBJECT NAME:
TEST
DOB: 07/07/78
STATE/D.L.: MO/NA
ARRESTING OFFICER:
PREMITT
OFFICER I.D.: 981
TESTING OFFICER:
SRME

OFFICER I.D.: 981
PERMIT NUMBER: 220269
EXPIRATION DATE: 09/13/14
MISCELLANEOUS DATA:
TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	VERIFIED	14:43
INTERNAL STANDARD			14:43
RADIO INTERFERENCE			

Operator Signature *[Signature]*

Face This Side Down - This Edge In First

BAC DataMaster

Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 2044272
02/10/14
14:38

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP:
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUALITY STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde+ghijklmnop
qrstuvwxy{|}~

Operator Signature *A. J. [Signature]*

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



JEFFERY L PREWITT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/13/2012

Number 220269

Expires 09/13/2014

Director of State Public Health Laboratory

Director, Department of Health