



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 11:54 am, Oct 15, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204177	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 10/02/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Troop D Satellite, Carthage, Missouri		TIME OF INSPECTION 11:15 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed value where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>10-02-14 11:15pm</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES LOT # 13280 EXP. DATE 10/06/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G6754 EXP. DATE 01/28/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ← .101	TEST 2 ← .101	TEST 3 ← .101
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This unit is operating within the rules and regulations of the Missouri Department of Health and Senior Services.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Jeffery L. Prewitt
TYPE II PERMIT NUMBER/EXPIRATION DATE 240067 03/07/2016	TELEPHONE NUMBER (417) 895-6868

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

JEFFERY L PREWITT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240067

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 460 (2/71) (6-10)

LAB 4 (R5-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **PREWITT, JEFFERY**
Permit No **240067**
Date Issued **3/7/2014** Date Expires **3/7/2016**

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
10/02/14

TESTING OFFICER:
PREMITT/J/L
OFFICER I.D.: 981
PERMIT NUMBER: 240067
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:
TEST

--- SUPERVISOR NOTE ---

BLANK TEST	.000	23:30
INTERNAL STANDARD	VERIFIED	23:30
EXTERNAL STANDARD	.101	23:31
BLANK TEST	.000	23:32
EXTERNAL STANDARD	.101	23:32
BLANK TEST	.000	23:33
EXTERNAL STANDARD	.101	23:33
BLANK TEST	.000	23:34

n = 9
SIM = 1
AVG = .101

Operator Signature

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
10/02/14
23:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

REACTOR: 49C

SAMPLE CHAMBER: OKAY

FLOW DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUANT. STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

M1 AKLHMPORSTUWXYZLNJ^_ abcdefghijklmnop
0123456789<=>?@ABCDEFGHIJKL

Operator Signature

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204177
10/02/14

ARREST TIME: 20:00
SUBJECT NAME:
TEST

JOB: 09/25/77 SEX: M

STATE A.L.: MO/NONE

ARRESTING OFFICER:

FRANCY/L/L

OFFICER I.D.: 961

TESTING OFFICER:

NAME

OFFICER I.D.: 981

PERMIT NUMBER: 240067

EXPIRATION DATE: 03/07/16

MISCELLANEOUS DATA:

TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	23:26
INTERNAL STANDARD	VERIFIED	23:26
RAMO INTERFERENCE		

Operator Signature

