DOH

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 9:15 am, Jun 18, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
204184	Missouri State Highway Patrol	5-30-14
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
CEDAR COUNTY JAIL, STOCKTON, MO		2100

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 5-30-14 2100
- [x] COMPUTER — [x] DETECTOR
- [x] PROGRAM — [x] FILTERS
- [x] HEATERS SAMPLE CHAMBER 50 °C °C — [x] QUARTZ STANDARD
- [x] FLOW DETECTOR — [x] CALIBRATION
- [x] PUMP HIGH SPEED — [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER REPCO MARKETING INC LOT # 13001 EXP. DATE 3-7-15
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C °C SIMULATOR SN 66757 EXP. DATE 12-20-14
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛	TEST 2 ☛	TEST 3 ☛
.098	.098	.099

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
Ø	Ø	1	1	Ø	Ø

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within Missouri Department of Health and Senior Services specifications.

MOVED INSTRUMENT TO NEW JAIL FACILITY AT 202 S. HIGH, STOCKTON, MO.

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Tpr. D.W. Scotten	Trooper D.W. Scotten
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240071 03/07/2016	(417) 895-6868

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscrimatory basis

REPCO MARKETING INC.

3101-168 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5460

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **13001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1215** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **March 8, 2013** The expiration date for this lot number is **March 7, 2015** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

DOUGLAS W SCOTTEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240071

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SCOTTEN, DOUGLAS
Permit No 240071
Date Issued 3/7/2014 **Date Expires** 3/7/2016

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204184
05/30/14
21:08

---- DIAGNOSTIC CHECK ----

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 50c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

SIMULATOR STANDARD: OKAY

PARTITION: OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature TPO J.W. Scotten

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204184
05/30/14

ARREST TIME: 21:00
SUBJECT NAME:
DOE/JON
DOB: 01/01/90 SEX: M
STATE/D...: MO/12345
ARRESTING OFFICER:
SCOTTEN/D/W
OFFICER I.D.: 938
TESTING OFFICER:
SCOTTEN/D/W
99 OFFICER I.D.: 938
PERMIT NUMBER: 240071
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

---- BREATH ANALYSIS ---

BLANK TEST	.000	21:59
INTERNAL STANDARD	VERIFIED	21:00
RADIO INTERFERENCE		

Operator Signature TPO J.W. Scotten

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204164
05/30/14

TESTING OFFICER:
SCOTTEN/D/W
OFFICER I.D.: 938
PERMIT NUMBER: 240071
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

---- SUPERVISOR MODE ----

BLANK TEST	.000	21:22
INTERNAL STANDARD	VERIFIED	21:22
EXTERNAL STANDARD	.098	21:22
BLANK TEST	.000	21:23
EXTERNAL STANDARD	.098	21:24
BLANK TEST	.000	21:24
EXTERNAL STANDARD	.099	21:25
BLANK TEST	.000	21:25

N = 3
STM. = .1
AVG. = .0983

Operator Signature Trp. D.W. Scotten