

DCH



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

received 10/19/14 report #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
By Carol Day at 12:29 pm, Oct 28, 2014

DATAMASTER SN 204188	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 10-11-14
LOCATION OF INSTRUMENT (STREET AND CITY) ST. CLAIR COUNTY JAIL, OSCEOLA, MO		TIME OF INSPECTION 1937

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 10-11-14 1937
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50°C °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES INC LOT # 13290 EXP. DATE 10-29-15
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0°C °C SIMULATOR SN 66757 EXP. DATE 12-20-14

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 = .095	TEST 2 = .095	TEST 3 = .095
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 0
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within Missouri Department of Health and Senior Services specifications.

INSPECTING OFFICER	
SIGNATURE <i>Trooper D.W. Scotten</i>	PRINT FULL NAME Trooper D.W. Scotten
TYPE II PERMIT NUMBER/EXPIRATION DATE 240071 03/07/2016	TELEPHONE NUMBER (417) 895-6868

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901**



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

DOUGLAS W SCOTTEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240071

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator SCOTTEN, DOUGLAS
 Permit No 240071
 Date Issued 3/7/2014 Date Expires 3/7/2016

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204188
10/11/14


TESTING OFFICER:
SCOTTEN/D/W
OFFICER I.D.# 938
PERMIT NUMBER: 240071
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	19:48
INTERNAL STANDARD	VERIFIED	19:48
EXTERNAL STANDARD	.095	19:49
BLANK TEST	.000	19:49
EXTERNAL STANDARD	.095	19:50
BLANK TEST	.000	19:51
EXTERNAL STANDARD	.095	19:51
BLANK TEST	.000	19:52

N = 3
MIN. = .1
AVG. = .095

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**


MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204188
10/11/14

ARREST TIME: 19:00
SUBJECT NAME:
JOE/JOHN
DOB: 01/01/90 SEX: M
STATE/D.L.#: MO/12345
ARRESTING OFFICER:
SCOTTEN/D/W
OFFICER I.D.# 938
TESTING OFFICER:
SAME
OFFICER I.D.# 938
PERMIT NUMBER: 240071
EXPIRATION DATE: 03/07/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	19:56
INTERNAL STANDARD	VERIFIED	19:56
RADIO INTERFERENCE		

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204188
10/11/14
19137

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS:	
SAMPLE CHAMBER:	50c
FLOW DETECTOR:	OKAY
PUMP:	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUINCY STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

11/11/14 10:11:14 204188 19137
H I J K L P N O P Q R S T U V W X Y Z 1 2 3 4 5 6 7 8 9 0
a b c d e f g h i j k l m n o p q r s t u v w x y z 1 2 3 4 5 6 7 8 9 0

Operator Signature

