



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

RECEIVED 3/5/14-CD REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed in service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

**REVIEWED**  
 By Carol Day at 3:37 pm, Mar 31, 2014

DATAMASTER SN 204192	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/28/2014
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) McDonald County Jail, Pineville	TIME OF INSPECTION 10:08 pm
---	--------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 22:08
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES, INC. LOT # 13280 EXP. DATE 10/16/2015

<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3143 EXP. DATE 01/16/2015
--

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .098	TEST 2 .099	TEST 3 .099
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 1	(.10-.14) 4	(.15-.19) 4	OVER .19 2
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).  
 NO ALTERATION OR MODIFICATION WAS MADE TO THIS INSTRUMENT.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME Josey J. Long
TYPE II PERMIT NUMBER/EXPIRATION DATE 220366 10/16/2014	TELEPHONE NUMBER (417) 895-6868

RETURN COMPLETED REPORT TO THE:  
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

890 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-564-6470

## CERTIFICATE OF ANALYSIS

**Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATA MASTER SERIAL NUMBER 204192  
02/28/14

TESTING OFFICER:  
LONG JOSEY, J

TEST TIME: 21:00

PERMIT NUMBER: 280366

EXPIRATION DATE: 10/16/14

MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	22:11
INTERNAL STANDARD	VERIFIED	22:11
INTERNAL STANDARD	.099	22:11
BLANK TEST	.000	22:11
INTERNAL STANDARD	.099	22:11
INTERNAL STANDARD	.000	22:11
INTERNAL STANDARD	.099	22:11
INTERNAL STANDARD	.000	22:11
INTERNAL STANDARD	.099	22:11

Signature  
Printed on recycled paper with 50% post-consumer waste

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATA MASTER SERIAL NUMBER 204192  
02/28/14

TESTING OFFICER:  
LONG JOSEY, J

TEST TIME: 21:00

PERMIT NUMBER: 280366

EXPIRATION DATE: 10/16/14

MISCELLANEOUS DATA:

--- DIAGNOSTIC CHECK ---

COMPUTER	OKAY
PROGRAM	OKAY
HEATERS	OKAY
SAMPLE MANIPULATOR	OKAY
FLOW DETECTOR	OKAY
PUMP	OKAY
HIGH SPEED	OKAY
DETECTORS	OKAY
FILTERS	OKAY
QUARTZ STANDARD	OKAY
CALIBRATION	OKAY

PRINTER TEST  
OKAY

Signature  
Printed on recycled paper with 50% post-consumer waste

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATA MASTER SERIAL NUMBER 204192  
02/28/14

TESTING OFFICER:  
LONG JOSEY, J

TEST TIME: 21:00

PERMIT NUMBER: 280366

EXPIRATION DATE: 10/16/14

MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:11
INTERNAL STANDARD	VERIFIED	22:11
RADIO INTERFERENCE		

Signature  
Printed on recycled paper with 50% post-consumer waste

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



JOSEY J LONG

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

**DATAMASTER; ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 10/16/2012

Number 220366

Expires 10/16/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)