



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

RECEIVED 4/5/14-CD **REPORT #6**

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 25 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

**REVIEWED**  
 By Carol Day at 4:18 pm, Apr 23, 2014

DATAMASTER SN 204192	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 03/31/2014
LOCATION OF INSTRUMENT (STREET AND CITY) McDonald County Jail, Pineville		TIME OF INSPECTION 9:49 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 09:49
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES, INC. LOT # 13290 EXP. DATE 10/29/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3143 EXP. DATE 01/16/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .098	TEST 2  .099	TEST 3  .099
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	2	(.10-.14)	1	(.15-.19)	4	OVER .19	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

NO ALTERATION OR MODIFICATION WAS MADE TO THIS INSTRUMENT.

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME Josey J. Long
TYPE II PERMIT NUMBER/EXPIRATION DATE 240062 03/07/2016	TELEPHONE NUMBER (417) 895-6868

RETURN COMPLETED REPORT TO THE:  
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901



## GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204192  
03/31/14  
09:49

DIAGNOSTIC CHECK ---

INTER: OKAY  
RAM (04-07-2009): OKAY  
ERS  
LE CHAMBER: 50c  
DETECTOR: OKAY  
SPEED: OKAY  
CTOR: OKAY  
ERS: OKAY  
TZ STANDARD: OKAY  
BRATION: OKAY

PRINTER TEST  
\$%()\*+,-./0123456789:;<=>@ABCDEFGHI  
LMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
tuvwxyz{|}~

ture



Operator Signature

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204192  
03/31/14

TESTING OFFICER:  
LONG/J/J  
OFFICER I.D.: 1282  
PERMIT NUMBER: 240062  
EXPIRATION DATE: 03/07/16  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000  
INTERNAL STANDARD VERIFIED  
EXTERNAL STANDARD .098  
BLANK TEST .000  
EXTERNAL STANDARD .099  
BLANK TEST .000  
EXTERNAL STANDARD .099  
BLANK TEST .000

N = 3  
SIM. = .1  
AVG. = .0986

09:53  
09:53  
09:53  
09:54  
09:54  
09:55  
09:55  
09:56

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket


MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204192  
03/31/14

ARREST TIME: 09:00  
SUBJECT NAME:  
TEST  
DOB: 01/01/01 SEX: M  
STATE/D.L.: MO/  
ARRESTING OFFICER:  
LONG/J/J  
OFFICER I.D.: 1282  
TESTING OFFICER:  
LONG/J/J  
OFFICER I.D.: 1282  
PERMIT NUMBER: 240062  
EXPIRATION DATE: 03/07/16  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000  
INTERNAL STANDARD VERIFIED  
RADIO INTERFERENCE

09:59  
09:59



Operator Signature



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**JOSEY J LONG**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240062

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-593-0771 (6-10)

LAB 4 (PG-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator LONG, JOSEY  
 Permit No 240062  
 Date Issued 3/7/2014 Date Expires 3/7/2016