



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED** REPORT #6  
 By Carol Day at 11:52 am, Jun 02, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 3 months).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204192	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 05/26/2014
LOCATION OF INSTRUMENT (STREET AND CITY) McDonald County Jail, Pineville		TIME OF INSPECTION 8:45 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 05/26/14 08:45
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES, INC. LOT # 13290 EXP. DATE 10/29/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3143 EXP. DATE 01/16/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .097	TEST 2  .098	TEST 3  .098
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 2	(0-.04) 9	(.05-.09) 2	(.10-.14) 5	(.15-.19) 3	OVER .19 4
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).  
 NO ALTERATION OR MODIFICATION WAS MADE TO THIS INSTRUMENT.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME Josey J. Long
TYPE II PERMIT NUMBER/EXPIRATION DATE 240062 03/07/2016	TELEPHONE NUMBER (417) 895-6868

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*


# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204192  
05/26/14  
08345

### --- DIAGNOSTIC CHECK ---

PUTER: OKAY  
BRAIN (04-07-2009): OKAY  
THERMIST: OKAY  
SAMPLE CHAMBER: 50c  
WAVELENGTH DETECTOR: OKAY  
PUMP SPEED: OKAY  
ELECTROLYTE: OKAY  
THERMIST: OKAY  
BRIEFLING STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
\*%#&@!~<0123456789:;{<=>0123456789:;  
KLMNOPQRSTUVWXYZ\_abcdefghijklmnopqrstuvwxyz  
stuvwxz(1)~>

Operator Signature   
Printed on recycled paper with agr-based ink.

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204192  
05/26/14


### TESTING OFFICER:

LONG/JOSEY/J  
OFFICER I.D.: 1282  
PERMIT NUMBER: 240052  
EXPIRATION DATE: 03/07/16  
MISCELLANEOUS DATA:

### --- SUPERVISOR MODE ---

BLANK TEST .000 08:50  
INTERNAL STANDARD .097 08:51  
EXTERNAL STANDARD .000 08:51  
BLANK TEST .098 08:52  
EXTERNAL STANDARD .000 08:52  
INTERNAL STANDARD .098 08:53  
EXTERNAL STANDARD .000 08:53

N = 3  
SIM. = .1  
AVG. = .0976

Operator Signature   
Printed on recycled paper with agr-based ink.

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204192  
05/26/14

### ARREST TIME: 08:30

SUBJECT NAME: TEST  
JOB: 01/01/01 SEX: M  
STATE/D.L.: MO/  
ARRESTING OFFICER:  
LONG/JKJ  
OFFICER I.D.: 1282  
TESTING OFFICER:  
LONG/JKJ  
OFFICER I.D.: 1282  
PERMIT NUMBER: 240052  
EXPIRATION DATE: 03/07/16  
MISCELLANEOUS DATA:

### --- BREATH ANALYSIS ---

BLANK TEST .000 08:57  
INTERNAL STANDARD .000 08:57  
RADIO INTERFERENCE

Operator Signature   
Printed on recycled paper with agr-based ink.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**JOSEY J LONG**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240062

EXPIRES 3/7/2016

MO.663-DZ.71-(6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4.(R6-10)



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator LONG, JOSEY  
Permit No 240062  
Date Issued 3/7/2014 Date Expires 3/7/2016