



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

RECEIVED

REPORT #1

By Carol Day at 9:47 am, Sep 05, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN 500072	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 09/04/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 2920 North Shamrock Road, Jefferson City		TIME OF INSPECTION 09:17:55

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC RECORD	
DATE AND TIME <u>09/04/2014 09:17:57</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.8°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>48.1°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS	
<input type="checkbox"/> SIMULATOR STANDARD	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE

<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u>	LOT # <u>AG402702</u>	EXP. DATE <u>01/27/2016</u>
--	------------------------------	------------------------------------

<input type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) _____	SIMULATOR SN _____	SIMULATOR EXP DATE _____
---	---------------------------	---------------------------------

<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.		
<input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE		
<input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE		
<input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE		

TEST 1: 0.077	TEST 2: 0.077	TEST 3: 0.077
----------------------	----------------------	----------------------

<input checked="" type="checkbox"/> PERFORM R.F.I. TEST
--

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:					
REFUSALS: 0	0-.04: 0	.05-.09: 0	.10-.14: 0	.15-.19: 0	OVER .19: 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

Performed instrument calibration.

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME DEWAYNE D CARVER
TYPE II PERMIT NUMBER 230101	EXPIRATION DATE 05/28/2015
	TELEPHONE NUMBER 573-751-4722

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, MO Department of Health and Senior Services
Southeast District Office
2875 James Blvd, Poplar Bluff, MO 63901



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 28-Jan-2014

Lot # AG402702

Exp. Date

27-Jan-2016

Cyl. Type

108

Component

Ethanol
Nitrogen

Certified Concentration

0.080 ± 0.002 BrAC (208 ppm)
Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.

EB0010581
EB0010570
EB0010285
EB0010561
EB0010681

Concentration

391.8 ppm
259.8 ppm
209.0 ppm
103.7 ppm
52.22 ppm

Serial No.

EB0010603
EB0010559
EB0010595
EB0010562
EB0010579

Concentration

392.5 ppm
258.9 ppm
208.9 ppm
104.9 ppm
52.94 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2014.08.11 15:13:08 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

DEWAYNE D CARVER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER, DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 05/28/2013
NUMBER 230101
EXPIRES 05/28/2015

W. S. ...

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paul Veeberling
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4, (R6-10)

MO 580-0771 (6-10)