



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**INTOX DMT MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 9:41 am, Sep 17, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN <b>500124</b>	NAME OF AGENCY <b>Missouri State Highway Patrol</b>	DATE OF INSPECTION <b>09/16/2014</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>2920 North Shamrock Road, Jefferson City</b>		TIME OF INSPECTION <b>14:04:52</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

**DIAGNOSTIC RECORD**

DATE AND TIME <u>09/16/2014 14:04:54</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.8°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>48.0°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

**BREATH ANALYZER ACCURACY STANDARDS**

SIMULATOR STANDARD       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS      LOT # AG418902      EXP. DATE 07/08/2016

SIMULATOR TEMP (34°C ± 0.2°C) \_\_\_\_\_      SIMULATOR SN \_\_\_\_\_      SIMULATOR EXP DATE \_\_\_\_\_

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
 Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: **0.099**      TEST 2: **0.099**      TEST 3: **0.099**

**PERFORM R.F.I. TEST**

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:**

REFUSALS: 0	0-.04: 0	.05-.09: 0	.10-.14: 0	.15-.19: 0	OVER .19: 0
-------------	----------	------------	------------	------------	-------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

Adjusted stylus sensitivity & positioning on LCD. Calibrated instrument.

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME <b>DEWAYNE D CARVER</b>
---------------	--

TYPE II PERMIT NUMBER <b>230101</b>	EXPIRATION DATE <b>05/28/2015</b>	TELEPHONE NUMBER <b>573-751-4722</b>
--	--------------------------------------	---

RETURN COMPLETED REPORT TO THE **Breath Alcohol Program, MO Department of Health and Senior Services  
 Southeast District Office  
 2875 James Blvd, Poplar Bluff, MO 63901**



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 8-Jul-2014

**Lot # AG418902**

**Exp. Date**

8-Jul-2016

**Cyl. Type**

108

**Component**

Ethanol  
Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (260 ppm)  
Balance

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

**Serial No.**

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

**Concentration**

391.8 ppm

259.8 ppm

209.0 ppm

103.7 ppm

52.22 ppm

**Serial No.**

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

**Concentration**

392.5 ppm

258.9 ppm

208.9 ppm

104.9 ppm

52.94 ppm

**Analytical Method:**

NDIR

Digitally signed by Quality Control  
Date: 2014.07.08 16:23:57 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

**Analyst:**

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**DEWAYNE D CARVER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV W/PRINTER, DATAMASTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 05/28/2013  
NUMBER 230101  
EXPIRES 05/28/2015

*W. S. ...*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Earl ...*  
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 580-0771 (6-10)