



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
CMI INTOXILYZER 5000 MAINTENANCE REPORT

RECEIVED
 By Carol Day at 8:59 am, Jan 13, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 36 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOXILYZER 5000 SN 660002504	NAME OF AGENCY Oronogo Police Department	DATE OF INSPECTION 01/08/2014
----------------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 653 E. Central Oronogo MO 64855	TIME OF INSPECTION 1:23 am
---	-------------------------------

CHECKLIST: Place a mark by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DVM TEST: (.350 ± .150) _____ 423

DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (FROM PRINTOUT) 01-08-2014 01:23

CHARACTER DISPLAY TEST

PRINT TEST (PRINTOUT ATTACHED)

SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 13280 EXP. DATE 10/16/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD1616 EXP. DATE 07/10/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .100	TEST 2 → .103	TEST 3 → .100
---------------	---------------	---------------

PERFORM RFI TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	0-04	0	.05-.09	0	.10-.14	0	.15-.19	0	Over .19	0
----------	---	------	---	---------	---	---------	---	---------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Seth Talbott
---------------	---------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230336 12-23-2015	TELEPHONE NUMBER (417) 673-1916
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, Missouri Department of Health and Senior Services
 Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

**GUTH LABORATORIES, INC.**

620 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

THIS PAGE IS THE ORIGINAL FORM NUMBER 019010

THIS PAGE IS THE ORIGINAL FORM NUMBER 046010

SI 66-082504
E735-23
INHALID TEST
INHIBITED - RFI

01/08/2014
01:32

ORANGE POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
MO MODEL 3000 SI 66-082504
01/08/2014

TEST	BAC	TIME
AIR BLANK	.000	01:22
CAL. CHECK	.100	01:27
AIR BLANK	.000	01:28
CAL. CHECK	.100	01:28
AIR BLANK	.000	01:28
CAL. CHECK	.100	01:29
AIR BLANK	.000	01:29

NO RFI PRESENT

SUBJECT'S NAME

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

TIME FIRST OBSERVED

INSTRUMENT LOCATION

OFFICER

ADDITIONAL INFORMATION AND/OR REMARKS

OFFICER

ADDITIONAL INFORMATION AND/OR REMARKS



SN 68-002504 01/09/2014
E735, 23 01:24

ABCDEFGHIJKLMNORSTUVWXYZ0123
ABCDEFGHIJKLMNORSTUVWXYZ0123456789
ABCDEFGHIJKLMNORSTUVWXYZ0123456789
ABCDEFGHIJKLMNORSTUVWXYZ0123456789
ABCDEFGHIJKLMNORSTUVWXYZ0123456789
ABCDEFGHIJKLMNORSTUVWXYZ0123456789
ABCDEFGHIJKLMNORSTUVWXYZ0123456789
ABCDEFGHIJKLMNORSTUVWXYZ0123456789

BRANDS POLICE DEPARTMENT
DRIVEN BY ALCOHOL ANALYZER
NO. 1031L 8800 SN 68-002504
01/09/2014

DIAGNOSTIC TEST 01:23
FRESH CHECK E735, 23 PASSED
BATT CHECK PASSED
TEMP CHECK PASSED
PROXIMITY CHECK
BYPASS TEST PASSED
BYPASS SPEED PASSED
REF STABILITY PASSED
PDS STABILITY PASSED
REF RANGE PASSED
DIAGNOSTIC PASSED
PRINTER CHECK
ABCDEFGHIJKLMNORSTUVWXYZ
0123456789

DRIVER'S NAME

SUBJECT'S NAME

~~TO: WISCONSIN III~~
~~INSTRUMENT NO: 8800~~

~~INSTRUMENT NO: 8800~~
~~INSTRUMENT NO: 8800~~





STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
SETH TALBOTT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/23/2013
NUMBER 230336
EXPIRES 12/23/2015



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
Gal Vesterby

, acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO:580-0771 (18-10)

LAB-4 (R6.10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **TALBOTT, SETH**
Permit No **230336**
Date Issued **12/23/2013** Date Expires **12/23/2015**