



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
CMI INTOXILYZER 5000 MAINTENANCE REPORT

RECEIVED
By Carol Day at 9:33 am, May 27, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOXILYZER 5000 SN <u>66604934</u>	NAME OF AGENCY <u>Platte Woods Police Dept</u>	DATE OF INSPECTION <u>05-27-14</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>6750 N.W. Tower Drive Platte Woods</u>		TIME OF INSPECTION <u>0838</u>

CHECKLIST: Place a mark by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DVM TEST: (.350 ± .150) .413

DIAGNOSTIC CHECK (PRINTOUT ATTACHED) Passed DATE AND TIME (FROM PRINTOUT) 05-27-2014 / 08:41

CHARACTER DISPLAY TEST Passed

PRINT TEST (PRINTOUT ATTACHED) Passed

SIMULATOR SOLUTION SUPPLIER Cuth LOT # 13296 EXP. DATE 10-29-2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIMULATOR SN SD1631 EXP. DATE 05-05-2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <u>.098</u>	TEST 2 <u>.099</u>	TEST 3 <u>.098</u>
--------------------	--------------------	--------------------

PERFORM RFI TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	0-04 <u>0</u>	05-09 <u>0</u>	10-14 <u>0</u>	15-19 <u>0</u>	Over .19 <u>0</u>
-------------------	---------------	----------------	----------------	----------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <u>Richard Lengley</u>	PRINT FULL NAME <u>Richard Lengley</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>2801981 / Exp 08-13-2014</u>	TELEPHONE NUMBER <u>816-741-6688</u>

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, Missouri Department of Health and Senior Services
Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

**GUTH LABORATORIES, INC.**

890 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

THIS SIDE UP, THIS EDGE IN. FORM NUMBER 015010

THIS SIDE UP, THIS EDGE IN. FORM NUMBER 015010

SN 66-004934
E735.23
INVALID TEST
INHIBITED - RFI

05/27/2014
08:47

PLATTE WOODS POLICE DEPT.
INTOXILYZER - ALCOHOL ANALYZER
NO MODEL 5000 SN 66-004934
05/27/2014

TEST	%BAC	TIME
AIR BLANK	.000	08:44
CAL. CHECK	.033	08:44
AIR BLANK	.000	08:44
CAL. CHECK	.033	08:45
AIR BLANK	.000	08:45
CAL. CHECK	.033	08:46
AIR BLANK	.000	08:46

NO RFI PRESENT

SUBJECT'S NAME

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

TIME FIRST OBSERVED

INSTRUMENT LOCATION

Richard Jeffrey
OPERATOR

Richard Jeffrey
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

ADDITIONAL INFORMATION AND/OR REMARKS



State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



RICHARD LANGLEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 08/13/2012

Number 220198

Expires 08/13/2014

Director of State Public Health Laboratory

Director, Department of Health