

BRIDGETON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
NO MODEL 5000 SN 66-005164
03/06/2014

SN 66-005164 03/06/2014
E735.23 23:36

DIAGNOSTIC TEST 23:35

PROM CHECK E735.23 PASSED
RAM CHECK PASSED
TEMP CHECK PASSED
PROCESSOR CHECK
SYNC PULSE PASSED
SYNC SPEED PASSED
NEG STABILITY PASSED
POS STABILITY PASSED
REF RANGE PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMN0PQRSTUVWXYZ
0123456789

ABCDEFGHIJKLMN0PQRSTUVWXYZ0123
ABCDEFGHIJKLMN0PQRSTUVWXYZ0123456789
ABCDEFGHIJKLMN0PQRSTUVWXYZ0123456789!@#
ABCDEFGHIJKLMN0PQRSTUVWXYZ0123456789!@#
ABCDEFGHIJKLMN0PQRSTUVWXYZ0123456789!@#
ABCDEFGHIJKLMN0PQRSTUVWXYZ0123456789!@#
ABCDEFGHIJKLMN0PQRSTUVWXYZ0123456789!@#

SUBJECT'S NAME

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

TIME FIRST OBSERVED

INSTRUMENT LOCATION

OPERATOR

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

SN 66-005164
E735.23
INVALID TEST
INHIBITED - RFI

03/06/2014
23:45

BRIDGETON POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
NO MODEL 5000 SN 66-005164
03/06/2014

TEST	%BAC	TIME
AIR BLANK	.000	23:40
CAL. CHECK	.097	23:41
AIR BLANK	.000	23:41
CAL. CHECK	.096	23:42
AIR BLANK	.000	23:42
CAL. CHECK	.096	23:42
AIR BLANK	.000	23:43

NO RFI PRESENT

SUBJECT'S NAME

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

TIME FIRST OBSERVED

INSTRUMENT LOCATION

KWAAP 201 [Signature]
OPERATOR

KWAAP 201 [Signature]
OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).


The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013

The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
DAVID C KNAPP

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/26/2013

NUMBER 230177

EXPIRES 8/26/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KNAPP, DAVID
 Permit No 230177
 Date Issued 8/26/2013 Date Expires 8/26/2015