



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**CMI INTOXILYZER 5000 MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 9:44 am, Jul 11, 2014

REPORT #4

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOXILYZER 5000 SN 66-005179	NAME OF AGENCY Raytown Police Department	DATE OF INSPECTION 07/09/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 10000 E 59 Street Raytown, MO 64133		TIME OF INSPECTION 10:27 am

**CHECKLIST:** Place a mark by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DVM TEST: (.350 ± .150) \_\_\_\_\_ .341

DIAGNOSTIC CHECK (PRINTOUT ATTACHED) \_\_\_\_\_ DATE AND TIME (FROM PRINTOUT) 07/09/2014 10:27

CHARACTER DISPLAY TEST

PRINT TEST (PRINTOUT ATTACHED )

SIMULATOR SOLUTION SUPPLIER Guth \_\_\_\_\_ LOT # 14030 \_\_\_\_\_ EXP. DATE 01/20/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ 34.0C \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ 2307 \_\_\_\_\_ EXP. DATE 07/10/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .102	TEST 2  .104	TEST 3  .103
--------------	--------------	--------------

PERFORM RFI TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	6	0-04	0	.05-.09	0	.10-.14	0	.15-.19	0	Over .19	0
----------	---	------	---	---------	---	---------	---	---------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Monthly maintenance was performed and the instrument was found to operating within Missouri Department of Health and Senior Services guidelines.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME Brian K. Porch
TYPE II PERMIT NUMBER/EXPIRATION DATE 220291 09/17/2014	TELEPHONE NUMBER (816) 737-6112

**RETURN COMPLETED REPORT TO THE:** Breath Alcohol Program, Missouri Department of Health and Senior Services  
 Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-684-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14030** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L**  $\pm$  3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



BRIAN K PORCH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/17/2012

Number 220291

Expires 09/17/2014

Director of State Public Health Laboratory

Director, Department of Health