



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
CMI INTOXILYZER 5000 MAINTENANCE REPORT

RECEIVED
 By Carol Day at 4:05 pm, Jan 15, 2014 #4

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOXILYZER 5000 SN 66-005220	NAME OF AGENCY Harrisonville PD	DATE OF INSPECTION 01/13/2014
----------------------------------	------------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 208 E. Pearl Harrisonville	TIME OF INSPECTION 12:28 pm
--	--------------------------------

CHECKLIST: Place a mark by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DVM TEST: (.350 ± .150) _____ .246
<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) _____ DATE AND TIME (FROM PRINTOUT) 01/13/2014 12:28
<input checked="" type="checkbox"/> CHARACTER DISPLAY TEST
<input checked="" type="checkbox"/> PRINT TEST (PRINTOUT ATTACHED)
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Repeco Marketing _____ LOT # 13002 _____ EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ 34.0 _____ SIMULATOR SN SD2254 _____ EXP. DATE 01/13/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ➔ .099	TEST 2 ➔ .100	TEST 3 ➔ .102
---------------	---------------	---------------

PERFORM RFI TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	0-0.04	0	0.05-0.09	0	0.10-0.14	0	0.15-0.19	0	Over .19	0
----------	---	--------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE ➔	PRINT FULL NAME Michael Davis
TYPE II PERMIT NUMBER/EXPIRATION DATE 230250 10/31/2015	TELEPHONE NUMBER (816) 380-8940

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, Missouri Department of Health and Senior Services
 Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

SN 66-005220 01/13/2014
E735.23 12:28

SN 66-005220 01/13/2014
E735.23 12:34
INVALID TEST
INHIBITED - RFI

ABCDEFGHIJKLMN OPQRSTUVWXYZ0123
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789!
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789!
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789!
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789!
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789!
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789!

HARRISONVILLE POLICE DEPT.
INTOXILYZER - ALCOHOL ANALYZER
NO MODEL 5000 SN 66-005220
01/13/2014

HARRISONVILLE POLICE DEPT.
INTOXILYZER - ALCOHOL ANALYZER
NO MODEL 5000 SN 66-005220
01/13/2014

TEST	%BAC	TIME
AIR BLANK	.000	12:30
CAL. CHECK	.099	12:31
AIR BLANK	.000	12:31
CAL. CHECK	.100	12:31
AIR BLANK	.000	12:32
CAL. CHECK	.102	12:32
AIR BLANK	.000	12:32

NO RFI PRESENT

DIAGNOSTIC TEST 12:28

PROM CHECK	E735.23	PASSED
RAM CHECK		PASSED
TEMP CHECK		PASSED
PROCESSOR CHECK		PASSED
SYNC PULSE		PASSED
SYNC SPEED		PASSED
NEG STABILITY		PASSED
POS STABILITY		PASSED
REF RANGE		PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMN OPQRSTUVWXYZ
0123456789

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

MICHAEL L DAVIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/31/2013

NUMBER 230250

EXPIRES 10/31/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator DAVIS, MICHAEL
Permit No 230250
Date Issued 10/31/2013 Date Expires 10/31/2015