





P E R M I T  
TYPE II



BRENT WORTHLEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 10/10/2012

Number 220358

Expires 10/10/2014

Director of State Public Health Laboratory

Director, Department of Health



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

THIS SIDE UP ↑ THIS EDGE IN

THIS SIDE UP ↑ THIS EDGE IN

01 24 000000  
 0351 24  
 00/27/2014  
 07:00  
 RAYMORE POLICE DEPARTMENT  
 INTERNAL TEST - ALLOCATION ANALYSIS  
 PRO MODEL 5000 01 00 000000  
 00/27/2014

RAYMORE POLICE DEPARTMENT  
 INTERNAL TEST - ALLOCATION ANALYSIS  
 PRO MODEL 5000 01 00 000000  
 00/27/2014

TEST	START	TIME
WIR. BLANK	0000	00:01
CAL. CHECK	0000	00:01
WIR. BLANK	0000	00:01
CAL. CHECK	0000	00:00
WIR. BLANK	0000	00:00
CAL. CHECK	0000	00:00
WIR. BLANK	0000	00:00

NO TEST PRESENT

THIS SIDE UP ↑ THIS EDGE IN

THIS SIDE UP ↑ THIS EDGE IN

01 24 000000  
 0351 24  
 00/27/2014  
 07:00  
 RAYMORE POLICE DEPARTMENT  
 INTERNAL TEST - ALLOCATION ANALYSIS  
 PRO MODEL 5000 01 00 000000  
 00/27/2014

RAYMORE POLICE DEPARTMENT  
 INTERNAL TEST - ALLOCATION ANALYSIS  
 PRO MODEL 5000 01 00 000000  
 00/27/2014

DYNAMIC TEST 07:57

PRIN. CHECK	07:56:20	PASSED
RAM CHECK		PASSED
TEMP. CHECK		PASSED
PROCESSOR CHECK		
SYNCH. PULSE		PASSED
SYNCH. SPEED		PASSED
MEM. STABILITY		PASSED
POE. STABILITY		PASSED
REF. RANGE		PASSED
DIAGNOSTIC		PASSED

PRINTER CHECK  
 RAYMORE POLICE DEPARTMENT  
 01 24 000000

**RAYMORE POLICE DEPARTMENT**  
 100 Municipal Circle  
 Raymore, Missouri 64083  
**(816) 331-0530**

**RAYMORE POLICE DEPARTMENT**  
 100 Municipal Circle  
 Raymore, Missouri 64083  
**(816) 331-0530**