



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14030** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

RAYMORE POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
NO MODEL 5000 SN 66-005283
08/22/2014

DIAGNOSTIC TEST 04:37

PROM CHECK E735.23 PASSED
RAM CHECK PASSED
TEMP CHECK PASSED
PROCESSOR CHECK
SYNC PULSE PASSED
SYNC SPEED PASSED
NEG STABILITY PASSED
POS STABILITY PASSED
REF RANGE PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMN OPQRSTUVWXYZ
0123456789

SN 66-005283 08/22/2014
E735.23 04:40

ABCDEFGHIJKLMN OPQRSTUVWXYZ0123
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789!@#*abcde
ABCDEFGHIJKLMN
ABCDEFGHIJKLMN OPQR
ABCDEFGHIJKLMN OPQRSTU
ABCDEFGHIJKLMN OPQRSTUVWXYZ0123456789!@#*abcde

Maintenance

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

Joshua B. Giaccone #919

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS



RAYMORE POLICE DEPARTMENT
INTOXILYZER - ALCOHOL ANALYZER
NO MODEL 5000 SN 66-005283
08/22/2014

TEST	%BAC	TIME
AIR BLANK	.000	04:42
CAL. CHECK	.102	04:42
AIR BLANK	.000	04:43
CAL. CHECK	.101	04:43
AIR BLANK	.000	04:43
CAL. CHECK	.102	04:44
AIR BLANK	.000	04:44

NO RFI PRESENT

SN 66-005283 08/22/2014
E735.23 04:50
INVALID TEST
INHIBITED - RFI

Maintenance

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

Joshua B Giaccone #919

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS



State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



JOSHUA GIACONE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 10/10/2012

Number 220343

Expires 10/10/2014

Director of State Public Health Laboratory

Director, Department of Health



Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations: 19CSR 25-30.051 (4).

Checked: 4/22/2014 Expires: 04/22/2015
Digital Therm. SN:094948
MSC Tech:DRL Temp:33.97
Agency: Raymore Police Dept
SD 2256



Technician Printed Name:

DAN LUCAS

Technician Signature:

[Handwritten Signature]

Date:

04/22/2014

Contact: Missouri Safety Center

Breath Alcohol Instrument Training Program

660-543-4834