





**GUTH LABORATORIES, INC.**

692 NORTH 87th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-664-5470

**CERTIFICATE OF ANALYSIS**

State of Missouri  
DEPARTMENT OF HEALTH



**PERMIT  
TYPE II**



MICHAEL DAVIS

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air, issued under the provisions of sections 577.020 through 577.041, RSMo 115.0.

exp. 09/07/2012  
no. 220232  
exp. 09/07/2014  
no. 00000000

Director of the State Health Laboratory  
Michael F. Danoff  
Director, Department of Health

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*Traceability:*  
This was conducted using Cerilliant Reference Standard lot number FN12211-02 whose values are traceable to NIST. Values are calibrated annually by an outside agency using NIST traceable weights. Verification is done prior to each use utilizing NIST traceable weights.

SN 66-005292  
E735.23  
INVALID TEST  
INHIBITED - RFI

02/28/2014  
12:31

BELTON POLICE DEPARTMENT  
INTOXILYZER - ALCOHOL ANALYZER  
MO MODEL 5000 SN 66-005292  
02/28/2014

TEST	REAC	TIME
AIR BLANK	.000	12:28
CAL. CHECK	.030	12:28
AIR BLANK	.000	12:28
CAL. CHECK	.057	12:29
AIR BLANK	.000	12:29
CAL. CHECK	.102	12:30
AIR BLANK	.000	12:30

NO RFI PRESENT

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

SN 66-005292  
E735.23

02/28/2014  
12:27

ABCDEFGHIJKLMNPOQRSTUVWXYZ0123  
ABCDEFGHIJKLMNPOQRSTUVWXYZ0123456789  
ABCDEFGHIJKLMNPOQRSTUVWXYZ012345678910#abcde  
ABCDEFGHIJKLMNPO  
ABCDEFGHIJKLMNPOQR  
ABCDEFGHIJKLMNPOQRSTU  
ABCDEFGHIJKLMNPOQRSTUVWXYZ012345678910#abcde

BELTON POLICE DEPARTMENT  
INTOXILYZER - ALCOHOL ANALYZER  
MO MODEL 5000 SN 66-005292  
02/28/2014

DIAGNOSTIC TEST 12:25

PRON CHECK	E735.23	PASSED
RAM CHECK		PASSED
TEMP CHECK		PASSED
PROCESSOR CHECK		PASSED
SYNC PULSE		PASSED
SYNC SPEED		PASSED
NEG STABILITY		PASSED
POS STABILITY		PASSED
REF RANGE	/	PASSED

DIAGNOSTIC PASSED

PRINTER CHECK  
ABCDEFGHIJKLMNPOQRSTUVWXYZ  
0123456789

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS