



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
 By Carol Day at 12:51 pm, Jun 11, 2014

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005838	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 06/06/2014	TIME OF INSPECTION 01:57
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	02:01	DRY	05514080	04/01/2016
Cal Check	0.080	02:01	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	02:02	N/A	N/A	N/A
Cal Check	0.081	02:02	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	02:03	0.080	CMI	
Cal Check	0.079	02:03	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	02:03	0.080		
Cal Check	0.079	02:03	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	02:03	0.081		
Pass			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.002	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Result	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	02:04
RAM Test	Pass		Subject Test	RFI*	02:04
EEPROM Checksum Test	Pass		Air Blank	0.000	02:05
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	00-04	05-09	10-14	15-19	OVER 19		
0	0	0	0	0	0	0	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT TESTED WITHIN DOHSS GUIDELINES

INSPECTING OFFICER

SIGNATURE 	PRINT NAME MAGERS	
TYPE II PERMIT NUMBER 240265	EXPIRATION DATE 06/02/2016	TELEPHONE NUMBER 8162345000