



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 8:31 am, Nov 10, 2014

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005842	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 11/04/2014	TIME OF INSPECTION 01:20
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 05514080A1	STANDARD EXPIRATION DATE 04/01/2016
Air Blank	0.000	01:23	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.079	01:23	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	01:24	CALIBRATION CHECK RESULT 1 0.079		
Cal Check	0.080	01:24	CALIBRATION CHECK RESULT 2 0.080		
Air Blank	0.000	01:25	CALIBRATION CHECK RESULT 3 0.081		
Cal Check	0.081	01:25	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%	SPREAD (MUST BE .005 OR LESS) 0.002	
Air Blank	0.000	01:26	<h1>Pass</h1>		

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	01:26
EEPROM Checksum Test	Pass		Subject Test	RFI*	01:26
Real Time Clock Test	Pass		Air Blank	0.000	01:27
DSP Test	Pass		<h1>Pass</h1>		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
1	5	1	2	5	4	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED WITHIN DHSS GUIDELINES

<b>INSPECTING OFFICER</b>		
SIGNATURE 	PRINT NAME NATHAN MAGERS	
TYPE II PERMIT NUMBER 240265	EXPIRATION DATE 06/02/2016	TELEPHONE NUMBER 8162345000