



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CFR 101.010 CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
By Carol Day at 12:51 pm, Jun 11, 2014

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005843	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 06/05/2014	TIME OF INSPECTION 22:48
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	22:55	DRY	05514080	04/01/2016
Cal Check	0.077	22:55	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	22:55	N/A	N/A	N/A
Cal Check	0.077	22:56	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	22:56	0.080	CMI	
Cal Check	0.077	22:57	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	22:57	0.077		
Cal Check	0.077	22:57	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	22:57	0.077		
Pass			CALIBRATION CHECK RESULT 3		
			0.077		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			3.7%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass				
EEPROM Checksum Test	Pass		Air Blank	0.000	22:58
Real Time Clock Test	Pass		Subject Test	RFI*	22:58
DSP Test	Pass		Air Blank	0.000	22:58
Analytical Stability Test	Pass		*RFI Detect		
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	3	4	8	4	3

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED WITHIN MISSOURI DOHSS GUIDLINES

INSPECTING OFFICER

SIGNATURE 	PRINT NAME MAGERS
---------------	----------------------

TYPE II PERMIT NUMBER 240265	EXPIRATION DATE 06/02/2016	TELEPHONE NUMBER 8162345000
---------------------------------	-------------------------------	--------------------------------