



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 8:32 am, Nov 10, 2014

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005841	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 11/04/2014	TIME OF INSPECTION 02:55
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 05514080A1	STANDARD EXPIRATION DATE 04/01/2016
Air Blank	0.000	02:57	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.079	02:58	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	02:58	CALIBRATION CHECK RESULT 1 0.079		
Cal Check	0.080	02:58	CALIBRATION CHECK RESULT 2 0.080		
Air Blank	0.000	02:59	CALIBRATION CHECK RESULT 3 0.080		
Cal Check	0.080	02:59	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%		
Air Blank	0.000	03:00	SPREAD (MUST BE .005 OR LESS) 0.001		

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		-----		
EEPROM Checksum Test	Pass		Air Blank	0.000	03:00
Real Time Clock Test	Pass		Subject Test	RFI*	03:01
DSP Test	Pass		Air Blank	0.000	03:01
Analytical Stability Test	Pass		*RFI Detect		
Modem Test	Pass		Pass		
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
1	0	1	0	1	2		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED WITHIN DHSS GUIDELINES

INSPECTING OFFICER			
SIGNATURE 	PRINT NAME NATHAN MAGERS		
TYPE II PERMIT NUMBER 240265	EXPIRATION DATE 06/02/2016	TELEPHONE NUMBER 8162345000	