



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
By Carol Day at 10:18 am, May 14, 2014

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005843	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 05/01/2014	TIME OF INSPECTION 23:18
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	23:21	DRY	05514080	04/01/2016
Cal Check	0.079	23:21	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	23:22	N/A	N/A	N/A
Cal Check	0.079	23:22	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	23:23	0.080	CMI	
Cal Check	0.079	23:23	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	23:23	0.079		
Cal Check	0.079	23:23	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	23:23	0.079		
Pass			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	23:24
RAM Test	Pass		Subject Test	RFI*	23:24
EEPROM Checksum Test	Pass		Air Blank	0.000	23:25
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	5	5	8	9	4	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED WITHIN THE MISSOURI DHSS GUIDLINES

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME TIM FILLPOT	
TYPE II PERMIT NUMBER 230158	EXPIRATION DATE 08/14/2015	TELEPHONE NUMBER 8164828195	



7 Burgin Dr. • P.O. Box 790 • Jacksonville, IL 62151-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 6358
Part #: BAC105L088T
Cylinder Size: 105L
Lot Number: 055140880A1
Expiration: 4/1/2016

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Component	Concentration:	Accuracy:	Method:
Ethanol	208 ppm	±1.00% or 3% BAC whichever is greater	NDIR
Nitrogen	balance		

NIIST Standard Reference Material
Cylinder No. CCI4290 / Job No. 09160202
Certified 1218 Impurity Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition
and direct sunlight. Do not allow storage area to
exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech

03/24/14
Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

ISO/IEC 17025:2005 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

TIMOTHY L. FILLPOT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/14/2013
NUMBER 230158
EXPIRES 8/14/2015
MO 048-071 (9-10)

[Signature]
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
MO-98-10

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The named candidate is authorized to operate an approved breath alcohol instrument for the determination of the alcoholic content of breath from or expired air in Missouri.

Operator: FILLPOT, TIMOTHY
Permit No. 230158
Date Issued 8/14/2013 Date Expires 8/14/2015