



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
By Carol Day at 8:22 am, Nov 10, 2014

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005844	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 11/04/2014	TIME OF INSPECTION 00:32
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 05514080A1	STANDARD EXPIRATION DATE 04/01/2016
Air Blank	0.000	00:34	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.079	00:34	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	00:35	CALIBRATION CHECK RESULT 1 0.079		
Cal Check	0.078	00:35	CALIBRATION CHECK RESULT 2 0.078		
Air Blank	0.000	00:36	CALIBRATION CHECK RESULT 3 0.079		
Cal Check	0.079	00:36	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.5%		
Air Blank	0.000	00:36	SPREAD (MUST BE .005 OR LESS) 0.001		

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	00:37
EEPROM Checksum Test	Pass		Subject Test	RFI*	00:37
Real Time Clock Test	Pass		Air Blank	0.000	00:38
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
1	1	0	0	0	1	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED WITHIN DHSS GUIDELINES

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME NATHAN MAGERS	
TYPE II PERMIT NUMBER 240265	EXPIRATION DATE 06/02/2016	TELEPHONE NUMBER 8162345000	