



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
 By Carol Day at 3:15 pm, Apr 11, 2014

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005847	LOCATION OF INSTRUMENT IPD DETENTION UNIT	DATE OF INSPECTION 04/09/2014	TIME OF INSPECTION 09:27
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	09:29	WET	13290	10/29/2015
Cal Check	0.098	09:30	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	09:31	34.0	DR4900	02/21/2015
Cal Check	0.099	09:31	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	09:32	0.100	GUTH LAB INC	
Cal Check	0.099	09:32	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	09:32	0.098		
Cal Check	0.099	09:32	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	09:33	0.099		
Pass			CALIBRATION CHECK RESULT 3		
			0.099		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.0%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	09:34
RAM Test	Pass		Subject Test	RFI*	09:34
EEPROM Checksum Test	Pass		Air Blank	0.000	09:34
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	31	2	1	1	5		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER			
SIGNATURE <i>Todd Hargis</i>		PRINT NAME TODD HARGIS	
TYPE II PERMIT NUMBER 230146	EXPIRATION DATE 08/01/2015	TELEPHONE NUMBER 816 325-7293	



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
TODD W HARGIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/1/2013

NUMBER 230146

EXPIRES 8/1/2015

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator HARGIS, TODD
Permit No 230146
Date Issued 8/1/2013 Date Expires 8/1/2015



Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations:
19 CSR 25-30.051 (4).

Checked: 2/21/2014 Expires: 02/21/2015
Digital Therm. SN:093752 Temp:34.00
MSC Tech:RW
Agency: INDEPENDENCE POLICE DEPT
DR4900




Technician Printed Name: ROBERT WELSH

Technician Signature: 

Date: 2/21/2014

Contact: Missouri Safety Center
Breath-Alcohol Instrument Training Program
660-543-4834

IN THE STATE OF MISSOURI
COUNTY OF JACKSON

AFFIDAVIT

Before me, the undersigned authority personally appears TOAD HAAGS.
Who, being by me duly sworn, deposed as follows:

My name is TOAD HAAGS I am of sound mind, capable of making this
affidavit, and personally acquainted with the facts herein stated.

I am the custodian of the records of *Intoxilyzer 5000* Serial number 80-005847.
Attached hereto are 4 pages of records from the **Independence Missouri Police
Department**. These pages of records are kept by the **Independence Missouri Police
Department** in regular course of business of the **Independence Missouri Police
Department** for an employee or representative of the **Independence Missouri Police
Department** with the knowledge of the act, event, condition, opinion, or diagnosis
recorded to make the record or to transmit information thereof to be included in such
record, and the record was made at or near the time of the act, event, condition, opinion
or diagnoses. The records attached hereto are the original or exact duplicates of the
original.

Toad Haags
AFFIANT

Subscribed and sworn to me on this 10 day of April, 2014.

Shari L. Rector
NOTARY PUBLIC

