



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED 7/24/14-CD

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
 By Carol Day at 3:46 pm, Aug 12, 2014

DATAMASTER SN 931097	NAME OF AGENCY Lawson Police Department	DATE OF INSPECTION 07-19-2014
LOCATION OF INSTRUMENT (STREET AND CITY) 104 W Third Street, Lawson MO 64062		TIME OF INSPECTION 1855

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07/19/14 18:56
<input checked="" type="checkbox"/> COMPUTER	<input type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **ReproMarketing** LOT # **13002** EXP. DATE **06-19-15**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0** °C SIMULATOR SN **DRS392** EXP. DATE **01/10/2015**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.098	TEST 2 0.098	TEST 3 0.098
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS —	(0-.04) —	(.05-.09) —	(.10-.14) 2	(.15-.19) —	OVER .19 —
------------	-----------	-------------	--------------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Conforms with DoHSS standards

INSPECTING OFFICER

SIGNATURE <i>Sgt Bruce Summa 201</i>	PRINT FULL NAME Sergeant Bruce Summa 201
---	--

TYPE II PERMIT NUMBER/EXPIRATION DATE 220319 09/28/2014	TELEPHONE NUMBER 816-580-7210
---	---

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901**

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097
07/19/14
18:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY

FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
q-rstuvwxyz{|}~

ERATOR SIGNATURE *Sgt Bruce Sumner 201*
ID STK # REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097
07/19/14

TESTING OFFICER:
SUMMA/BRUCE/A
OFFICER I.D.: 201
PERMIT NUMBER: 220319
EXPIRATION DATE: 09/28/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000 18:59
INTERNAL STANDARD VERIFIED 18:59
EXTERNAL STANDARD .098 19:00
BLANK TEST .000 19:00
EXTERNAL STANDARD .098 19:01
BLANK TEST .000 19:02
EXTERNAL STANDARD .098 19:02
BLANK TEST .000 19:02

N = 3
SIM. = .1
AVG. = .098

OPERATOR SIGNATURE *Sgt Bruce Sumner 201*
CARD STK # REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)
60036

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097
07/19/14

ARREST TIME: 18:00
SUBJECT NAME:
RFI/TEST
DOB: 11/12/45 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
SUMMA/BRUCE/A
OFFICER I.D.: 201
TESTING OFFICER:
SUMMA/BRUCE/A
OFFICER I.D.: 201
PERMIT NUMBER: 220319
EXPIRATION DATE: 09/28/14
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 19:06
INTERNAL STANDARD VERIFIED 19:06
RADIO INTERFERENCE

OPERATOR SIGNATURE *Sgt Bruce Sumner 201*
CARD STK # REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)
60036

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



BRUCE SUMMA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220319

Expires 09/28/2014

Director of State Public Health Laboratory

Director, Department of Health