



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED** REPORT #6  
 By Carol Day at 9:20 am, Aug 28, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 931097 NAME OF AGENCY Lawson Police Department DATE OF INSPECTION 08/19/2014

LOCATION OF INSTRUMENT (STREET AND CITY) 104 W Third Street Lawson MO 64062 TIME OF INSPECTION 0742

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 08/19/14 07:43

COMPUTER  DETECTOR

PROGRAM  FILTERS

HEATERS SAMPLE CHAMBER 49 °C  QUARTZ STANDARD

FLOW DETECTOR  CALIBRATION

PUMP HIGH SPEED  PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repro Marketing LOT # 13002 EXP. DATE 36-19-15

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN DR5392 EXP. DATE 01/10/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0/01 TEST 2 .100 TEST 3 .101

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 1 (0-.04) — (.05-.09) — (.10-.14) — (.15-.19) — OVER .19 —

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Conforms with DohSS Standards.

**INSPECTING OFFICER**

SIGNATURE Sgt Bruce Summa 201 PRINT FULL NAME Sgt Bruce Summa 201

TYPE II PERMIT NUMBER/EXPIRATION DATE 220319 09/28/2014 TELEPHONE NUMBER 816-580-7210

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**  
**LOT NUMBER: 13002**  
**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097  
08/19/14

ARREST TIME: 07:00  
SUBJECT NAME:  
RFI/TEST  
DOB: 11/12/34 SEX: M  
STATE/D.L.: MO/123456  
ARRESTING OFFICER:  
SUMMA/BRUCE/A  
OFFICER I.D.: 201  
TESTING OFFICER:  
SUMMA/BRUCE/A  
OFFICER I.D.: 201  
PERMIT NUMBER: 220319  
EXPIRATION DATE: 09/28/14  
MISCELLANEOUS DATA:  
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 07:55  
INTERNAL STANDARD VERIFIED 07:55  
RADIO INTERFERENCE

OPERATOR SIGNATURE *Sgt Bruce Summa 201*

ID STK # REORDER ALL SUPPLIES FROM N.P.A.S.  
2260 NORTH MAIN, MANSFIELD, OH 44003 419-526-6777 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097  
08/19/14  
07:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY

HEATERS  
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP  
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJK  
LMNOPQRSTUVWXYZ[\]^\_`abcdefshijklmnop  
q-rstuvwxyz{|}~

OPERATOR SIGNATURE *Sgt Bruce Summa 201*

CARD STK # REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097  
08/19/14

TESTING OFFICER:  
SUMMA/BRUCE/A  
OFFICER I.D.: 201  
PERMIT NUMBER: 220319  
EXPIRATION DATE: 09/28/14  
MISCELLANEOUS DATA:  
--- SUPERVISOR MODE ---

BLANK TEST .000 07:48  
INTERNAL STANDARD VERIFIED 07:48  
EXTERNAL STANDARD .101 07:48  
BLANK TEST .000 07:49  
EXTERNAL STANDARD .100 07:49  
BLANK TEST .000 07:50  
EXTERNAL STANDARD .101 07:50  
BLANK TEST .000 07:51

N = 3  
SIM. = .1  
AVG. = .1006

OPERATOR SIGNATURE *Sgt Bruce Summa 201*

CARD STK # REORDER ALL SUPPLIES FROM N.P.A.S.  
2260 NORTH MAIN, MANSFIELD, OH 44003 419-526-6727 (NPAS)

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



BRUCE SUMMA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220319

Expires 09/28/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)