



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 REPORT #6
 By Carol Day at 10:25 am, Dec 01, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 95 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 931697	NAME OF AGENCY Lawson Police Department	DATE OF INSPECTION 11-21-2014
LOCATION OF INSTRUMENT (STREET AND CITY) 104 W Third Street Lawson Mo 64062		TIME OF INSPECTION 0538

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11-21-2014 05:39
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **Repro Marketing** LOT # **14001** EXP. DATE **04-30-16**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0** °C SIMULATOR SN **DR5392** EXP. DATE **01/10/2015**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.100	TEST 2 0.101	TEST 3 0.102
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 1	(0-.04) —	(.05-.09) —	(.10-.14) —	(.15-.19) —	OVER .19 —
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Conforms with DOHSS standards

INSPECTING OFFICER

SIGNATURE Sgt Bruce Sumner 201	PRINT FULL NAME Sgt Bruce Sumner 201
TYPE / PERMIT NUMBER / EXPIRATION DATE 240362 10-03-2016	TELEPHONE NUMBER 816 580 7210

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901**

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 14001

EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.


This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014

The expiration date for this lot number is April 30, 2016 at

11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097
11/21/14

ARREST TIME: 05:00
SUBJECT NAME:
REF/TEST

DOB: 11/23/45 SEX: M
STATE/D.L.: MO/123456

ARRESTING OFFICER:
SUMMA/BRUCE/A

TESTING OFFICER:
SUMMA/BRUCE/A

OFFICER I.D.: 201
PERMIT NUMBER: 240062

EXPIRATION DATE: 10/03/16
MISCELLANEOUS DATA:

RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 05:51
INTERNAL STANDARD VERIFIED 05:51
RADIO INTERFERENCE

Operator Signature *Sgt. Bruce Summa 201*

2208-02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097
11/21/14

05:39

--- DIRMOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~++

Operator Signature *Sgt. Bruce Summa 201*

2208-02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 931097
11/21/14

TESTING OFFICER:
SUMMA/BRUCE/A

OFFICER I.D.: 201

PERMIT NUMBER: 240062

EXPIRATION DATE: 10/03/16

MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000 05:42
INTERNAL STANDARD VERIFIED 05:42
EXTERNAL STANDARD .100 05:43
BLANK TEST .000 05:43
EXTERNAL STANDARD .101 05:44
BLANK TEST .000 05:44
EXTERNAL STANDARD .102 05:45
BLANK TEST .000 05:45

N = 3
SIM. = .1
AVG. = .101

Operator Signature *Sgt. Bruce Summa 201*

2208-02



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2



PERMIT
TYPE II
BRUCE SUMMA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/3/2014
 NUMBER 240362
 EXPIRES 10/3/2016


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

 acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator SUMMA, BRUCE
 Permit No 240362
 Date Issued 10/3/2014 Date Expires 10/3/2016