MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 12:59 pm, Oct 07, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
940093	AVA POLICE DEPARTMENT	10/06/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
504 NW 12TH AVENUE AVA, MO. 65608 AVA POLICE DEPARTMENT		11:26 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 10/06/2014 11:26AM

☑ COMPUTER — ☑ DETECTOR

☑ PROGRAM — ☑ FILTERS

☑ HEATERS SAMPLE CHAMBER 48 °C — ☑ QUARTZ STANDARD

☑ FLOW DETECTOR — ☑ CALIBRATION

☑ PUMP HIGH SPEED — ☑ PRINTER

☑ INDICATOR LIGHTS

☑ SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES INC. LOT # 14110 EXP. DATE 05/01/2016

☑ SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD1617 EXP. DATE 03/17/2015

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
0.100	0.099	0.101

☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
1	0	0	0	0	2

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

THIS INSTRUMENT IS OPPERATING WITHIN DOH STANDARDS

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Lt. Trent Murray	LT. TRENT MURRAY
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
230236 10/17/2015	(417) 683-2931

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14110** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 5, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 1, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
AVA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940093
10/06/14

TESTING OFFICER:
MURRAY
OFFICER I.D.: 406
PERMIT NUMBER: 230236
EXPIRATION DATE: 10/17/15
MISCELLANEOUS DATA:
MONTHLY MAINT

--- SUPERVISOR MODE ---

BLANK TEST	.000	11:40
INTERNAL STANDARD	VERIFIED	11:40
EXTERNAL STANDARD	.100	11:40
BLANK TEST	.000	11:41
EXTERNAL STANDARD	.099	11:42
BLANK TEST	.000	11:42
EXTERNAL STANDARD	.101	11:43
BLANK TEST	.000	11:43

N = 3
SIM. = .1
AVG. = .1

Operator Signature

2208-02

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
AVA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940093
10/06/14

ARREST TIME: 11:00
SUBJECT NAME:
MURRAY
DOB: 04/22/78 SEX: M
STATE/D.L.: MO/1234567890
ARRESTING OFFICER:
MURRAY
OFFICER I.D.: 406
TESTING OFFICER:
MURRAY
OFFICER I.D.: 406
PERMIT NUMBER: 230236
EXPIRATION DATE: 10/17/15
MISCELLANEOUS DATA:
RFI TEST
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	11:30
INTERNAL STANDARD	VERIFIED	11:31
RADIO INTERFERENCE		

Operator Signature L.P. Scott Murray

2208-02

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
AVA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940093
10/06/14
11:26

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	48c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklm
pqrstuvwxyz{|}~€

Operator Signature L.P. Scott Murray

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

NORMAN T MURRAY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/17/2013

NUMBER 230236

EXPIRES 10/17/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

, acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri

Operator MURRAY, NORMAN

Permit No 230236

Date Issued 10/17/2013 Date Expires 10/17/2015

Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations: 19CSR 25-30.051 (4).

Checked: 03/17/2014 Expires: 03/17/2015
Digital Therm. SN:094948 Temp:34.00
MSC Tech:RW
Agency: AVA Police Dept.
SD1617

Technician Printed Name: ROBERT WELSH

Technician Signature:

Date: 3/17/2014

Contact: Missouri Safety Center

Breath Alcohol Instrument Training Program

660-543-4834