



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 2:46 pm, May 12, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|--|--|--------------------------------------|
| DATAMASTER SN 940097 | NAME OF AGENCY Sweet Springs Police Dept | DATE OF INSPECTION 5/01/14 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 324 S. Miller Sweet Springs | | TIME OF INSPECTION 23:07 |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

| | |
|--|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) | DATE AND TIME (from printout) 05/01/14 23:07 |
| <input checked="" type="checkbox"/> COMPUTER | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM 04/02/2009 | <input checked="" type="checkbox"/> FILTERS |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR | <input checked="" type="checkbox"/> CALIBRATION |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED | <input checked="" type="checkbox"/> PRINTER |

| |
|--|
| <input checked="" type="checkbox"/> INDICATOR LIGHTS |
| <input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Replco Marketing LOT # 12002 EXP. DATE 08/29/14 |
| <input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34°C ± 0.2° °C SIMULATOR SN SD3114 EXP. DATE NONE |

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

| | | |
|----------------------|----------------------|----------------------|
| TEST 1 = 100% | TEST 2 = 100% | TEST 3 = 100% |
|----------------------|----------------------|----------------------|

PERFORM R.F.I. TEST (PRINTOUT ATTACHED) **RFI Detected**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | |
|------------|-----------|-------------|-------------|-------------|------------|
| REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | OVER .19 0 |
|------------|-----------|-------------|-------------|-------------|------------|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within Department of Health Specifications using 100% solution from Replco Marketing bearing lot # 12002 manufacture date 08-30-2012 expiration date 08/29/2012

| | | |
|--|---|--|
| INSPECTING OFFICER | SIGNATURE <i>Melvin Taber</i> | PRINT FULL NAME Melvin Taber |
| TYPE I PERMIT NUMBER/EXPIRATION DATE 230211 10/02/15 | TELEPHONE NUMBER 660 335-6823 | |

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MELVIN L TABER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/2/2013

NUMBER 230211

EXPIRES 10/2/2015

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **TABER, MELVIN**
 Permit No **230211**
 Date issued **10/2/2013** Date Expires **10/2/2015**

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940897
05/01/14

TESTING OFFICER:
TABER MELVIN L
OFFICER I.D.# 902
PERMIT NUMBER: 030011
EXPIRATION DATE: 10-02-15
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE TEST
SUPERVISOR MODE

--- SUPERVISOR MODE ---

| | | |
|-------------------|----------|-------|
| BLANK TEST | .000 | 23:12 |
| INTERNAL STANDARD | VERIFIED | 23:12 |
| EXTERNAL STANDARD | .100 | 23:12 |
| BLANK TEST | .000 | 23:13 |
| EXTERNAL STANDARD | .100 | 23:13 |
| BLANK TEST | .000 | 23:14 |
| EXTERNAL STANDARD | .100 | 23:14 |
| BLANK TEST | .000 | 23:10 |

N = 3
SIM. = .009
AVG. = .1

Operator Signature

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940897
05/01/14
23:07

DIAGNOSTIC CHECK

| | |
|-----------------------|------|
| COMPUTER: | OKAY |
| PROGRAM (04-07 2009): | OKAY |
| HEATERS | |
| SAMPLE CHANGERS: | OK |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

! " # \$ % & ' () * + , - . / : ; < = > ? @ [\] ^ _ ` { | } ~ ! " # \$ % & ' () * + , - . / : ; < = > ? @ [\] ^ _ ` { | } ~

Operator Signature

2208-02

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940037
05/01/14

ARREST TIME: 02:50

SUBJECT NAME:

TABER/M.L.

DOB: 01/23/50 SEX: M

STATE/D.L.: MO/123456789

ARRESTING OFFICER:

TABER/MELVIN/L

OFFICER I.D.: 992

TESTING OFFICER:

TABER/MELVIN/L

OFFICER I.D.: 992

PERMIT NUMBER: 200011

EXPIRATION DATE: 10/08/15

MISCELLANEOUS DATA:

MONTHLY MAINTENANCE TEST

RADIO INTERFERENCE TEST

--- BREATH ANALYSIS ---

BLANK TEST

.000

03:01

INTERNAL STANDARD

VERIFIED

03:01

RADIO INTERFERENCE

Operator Signature

Melvin L. Taber