



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 12:14 pm, Sep 22, 2014
 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 940097	NAME OF AGENCY Sweet Springs PD	DATE OF INSPECTION 09/18/14
LOCATION OF INSTRUMENT (STREET AND CITY) 324 S. Miller Sweet Springs		TIME OF INSPECTION 20:15

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 09/18/14 20:15
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing LOT # 13002 EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34°C ± 0.2 °C SIMULATOR SN# 803144 EXP. DATE 08/25/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.099%	TEST 2 0.099%	TEST 3 0.099%
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED) **R F I detected**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within Dept of Health Specifications using 100% solution from RepCo marketing bearing lot #13002 manufacture date 06/20/2013 expiration date June 19/2015

INSPECTING OFFICER	SIGNATURE Melvin L. Taber	PRINT FULL NAME Melvin L. Taber
TYPE & PERMIT NUMBER/EXPIRATION DATE 230 211 10/02/2015	TELEPHONE NUMBER 660 335 6823	

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MELVIN L TABER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/2/2013

NUMBER 230211

EXPIRES 10/2/2015

MO 680-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **TABER, MELVIN**
 Permit No **230211**
 Date Issued **10/2/2013** Date Expires **10/2/2015**

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 948897
09/18/14

TESTING OFFICER:
TABER/MELVIN/L
OFFICER I.D. #: 902
PERMIT NUMBER: 230211
EXPIRATION DATE: 12-02-10
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE TESTS
SUPERVISOR: NONE

--- SUPERVISOR MODE ---

BLANK TEST	.000	20:18
INTERNAL STANDARD	VERIFIED	20:18
EXTERNAL STANDARD	.000	20:18
BLANK TEST	.000	20:19
EXTERNAL STANDARD	.000	20:20
BLANK TEST	.000	20:20
EXTERNAL STANDARD	.000	20:21
BLANK TEST	.000	20:21

N = 3
SIN. = .000
AVG. = .000

Operator Signature

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 948897
09/18/14
20:15

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (84-87-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTED TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n
o p q r s t u v w x y z () * +

Operator Signature

2208-02

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 948007
09/18/14

ARREST TIME: 19:40
SUBJECT NAME:
TRADER/MELVIN/L
DOB: 01/02/58 SEX: M
STATE D.L.L.: MO-123456789
ARRESTING OFFICER:
TRADER/MELVIN/L
OFFICER I.D.: 902
TESTING OFFICER:
TRADER/MELVIN/L
OFFICER I.D.: 902
PERMIT NUMBER: 230211
EXPIRATION DATE: 10/02/15
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE TEST
RADIO INTERFERENCE TEST

--- BREATH ANALYSIS ---

BLANK TEST	.800	20:25
INTERNAL STANDARD	VERIFIED	20:25
RADIO INTERFERENCE		

Operator Signature

