



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

RECEIVED  
 By Carol Day at 12:24 pm, Oct 28, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>940097</b>	NAME OF AGENCY <b>Sweet Springs P.D.</b>	DATE OF INSPECTION <b>10/18/14</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>324 S. Miller Sweet Springs</b>		TIME OF INSPECTION <b>12:48</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>10/18/14 12:48</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>49</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <b>RepCo Marketing</b> LOT # <b>13002</b> EXP. DATE <b>06/19/2015</b>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <b>34°C ± 0.2 °C</b> SIMULATOR SN <b>SD3144</b> EXP. DATE <b>08/25/15</b>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 = <b>1.079%</b>	TEST 2 = <b>1.079%</b>	TEST 3 = <b>1.079%</b>
------------------------	------------------------	------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

*This instrument is operating within Dept of Health specifications using 100% solution being lot # 13002 manufacture date 06/20/2013 and expiration date 06/19/2015.*

INSPECTING OFFICER	SIGNATURE <i>Melvin Taber</i>	PRINT FULL NAME <b>Melvin Taber</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>230211 10/02/15</b>	TELEPHONE NUMBER <b>660 335 6823</b>	

RETURN COMPLETED REPORT TO THE:  
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**  
**LOT NUMBER: 13002**  
**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**MELVIN L TABER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 10/2/2013

NUMBER 230211

EXPIRES 10/2/2015

MO 580-0771 (6-10)

  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES, acting director

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator TABER, MELVIN  
 Permit No 230211  
 Date issued 10/2/2013 Date Expires 10/2/2015

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940097  
10/18/14

ARREST TIME: 12:35  
SUBJECT NAME:  
TABER/M/L  
DOB: 01/23/53 SEX: M  
STATE D.L.: MO/122456789  
ARRESTING OFFICER:  
TABER/MELVIN/L  
OFFICER I.D.: 982  
TESTING OFFICER:  
TABER/M/L  
OFFICER I.D.: 982  
PERMIT NUMBER: 200011  
EXPIRATION DATE: 10-02-15  
MISCELLANEOUS DATA:  
MONTHLY MAINTENANCE TEST  
RADIO INTERFERENCE TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	12:09
INTERNAL STANDARD	VERIFIED	12:09
RADIO INTERFERENCE		

Operator Signature

2208-02

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940097  
10/18/14

TESTING OFFICER:  
TABER/MELVIN/L  
OFFICER I.D.: 982  
PERMIT NUMBER: 200011  
EXPIRATION DATE: 10-02-15  
MISCELLANEOUS DATA:  
MONTHLY MAINTENANCE TEST  
SUPERVISOR MODE

--- SUPERVISOR MODE ---

BLANK TEST	.000	12:04
INTERNAL STANDARD	VERIFIED	12:05
EXTERNAL STANDARD	.000	12:06
BLANK TEST	.000	12:06
EXTERNAL STANDARD	.000	12:06
BLANK TEST	.000	12:07
EXTERNAL STANDARD	.000	12:07
BLANK TEST	.000	12:08

N 2  
SIM. = .000  
AVG. = .000

Operator Signature

2208-02

